

Job Number/Description 940110 - VN202

Requested By JSL Date 01/10/94	Position MD
Location Vandagraph	Telephone No.

Description / Specification/Brief

Oxygen Analyser for diving
Waterproof, hand held, Clear display, Accuracy +/- 0.1% , Fuel cell based
with a method of physically measuring O₂ in diving cylinders

Tests to be carried out in-house.

Leading technical divers to be consulted. T. Mount IANTD. R. Palmer TDI, K Gurr IANTD. Diver
exhibition,

Client Trial Requirements

QC22