

Service Repair Sheet SRS67818

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lisa Vickers

Medivent

01-6280338

00007374

info@medivent.ie

06/May/2021

Robert Connor

Viamed

Quote

Notes 29/Apr/2021 Sarah Walton

29/Apr/2021 Sarah Walton

VM-2500-S

Serial number A1809150030

The Screen is black when its turned on.

05/May/2021 Catherine Spence

sent back by DPD

06/May/2021 Robert Connor

1 x VM2500S s/n A1809150030, with clear cover and li-poly battery.

Unit has been tampered with. Case screws removed from top, back label missing and 2 additional screws through back panel

Ready For quote

C. A. Neen 12.5.2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33670	Capnograph	4410520	A1809150030	N

Dist.

4440000 @ 0.5hr * @ 70 € per hr. half hour = € 35.

S/N, SLS, SRN

UPS x 1 @

32x2

23x15x15cm

2.5kg 1kg.



Tel: 01-628 0338 E-mail: info@medivent.ie


WEST YORKSHIRE BD20 7DT

DELIVERY/ SHIPPING NOTE

DATE: 4th May 2021

RECEIVED BY: _____

#NAME?



Select Currency

Pounds| Euros| Usd

Repair SRS67818 Values in Pounds @ 1: 1.00

SRN33670	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4440000	0.5	90.00	45.00
				Parts Cost	45.00

Dist €
 €70 per hr
 €35

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent