

DECLARATION OF CONFORMITY

APPLICATION OF MEDICAL DEVICE DIRECTIVE : 93-42/EEC (Annex II)
EN46001
ISO13485
ISO9001:2000

STANDARDS TO WHICH CONFORMITY IS DECLARED : ISO 7767
IEC/EN60601-1, IEC/EN60601-1-2
EN61000-4-2,3,4,6,8.

MANUFACTURER'S NAME : TELEDYNE INSTRUMENTS
ANALYTICAL INSTRUMENTS

MANUFACTURER'S ADDRESS : 16830 Chestnut Street
City of Industry, CA 91748-1020
U.S.A.

EQUIPMENT TYPE : Medical Device, Oxygen Analyzer & Monitor

EQUIPMENT CLASSIFICATION : Class IIa (Annex IX, Rule 10)

MODEL NUMBER : AX300 & MX300

THE UNDERSIGNED HEREBY DECLARE THAT THE EQUIPMENT
SPECIFIED ABOVE CONFORMS TO THE ABOVE DIRECTIVE(S) AND
STANDARD(S).

SIGNATURE: 

FULL NAME: Ray Khajavi

POSITION: Director, Quality Assurance

DATE: 02/26/03

PLACE: City of Industry, California