

Service Repair Sheet SRS67820

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Sandra Hart

Company/ Hospital Name

Priory Hospital Chelmsford

Department

Position

Direct Phone

General Phone

01245 345345

Opera Account

00012966

Email

sandrahart@priorygroup.com

Order Number

Date Received

12/May/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

For Calibration

Notes 04/May/2021 Zoey Teal

04/May/2021 Zoey Teal

Sending in a AlcroTrue for Calibration. Serial number 51100138

12/May/2021 Robert Connor

1 x alcotru s/n 51100138, with battery

Ready For quote

CGNEON 18-5-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33678	Accessory	4710010	51100138	

4780010 - changed by CG 18-5-2021  
~~4740010~~  
4780018 x 1 @ £25  
SIN, SRS, SRN.

NOW: 4730040 CG 19-5-2021  
~~4740010~~ x 1 @ £150  
New PCB Board.

UPS x 1 @ £10.