

Service Repair Sheet SRS67842

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Date Promised to Return: 01 Jan 1970 -3600
Nurul Amin
Bradford Royal Infirmary
Clinical Engineering
Specialist Healthcare Scientist
01274 364147
0127 454 2200
00000620
nurul.amin@bthft.nhs.uk

27/May/2021
Robert Connor
Viamed
Quote

Notes 25/May/2021 Zoey Teal

25/May/2021 Zoey Teal

Sending in a Tom Thumb. Serial number 0401193. The pressure will not go above 10 cmWG. They have replaced the pressure gauge with a new one but made no difference.

27/May/2021 Robert Connor

1 x tom thumb minus gauge s/n 0401193, tom thumb gauge

Ready For quote

CANEN 27-5-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33726	Tom Thumb	0310030	0401193	N

0380000 x 1 £90
S/N, SRS, SRN

0330212 x 1 £1.50
SRS, SRN

UPS x 1 £10.

Supplier's Detail's

Viamed Limited
FAO REPAIRS
15 Station Road
Cross Hills
Keighley
BD20 7DT

**Department of
Medical Physics**
Clinical Engineering
Bradford Royal Infirmary
Duckworth Lane
Bradford
West Yorkshire
BD9 6RJ
Tel: 01274 364147
Fax: 01274 364134

Purchase Order Number:

Date: 26/5/21

Please find enclosed the following item(s)

For Repair:	x
For Service Exchange	
For Replacement:	
For Calibration:	
For Service:	
Warranty Return:	

Quantity 1

Manufacturer: Viamed
Model: TOM THUMB
Type : Infant Resuscitaire

Serial Number: 0401193

Equipment Number: 56202
Returns Reference Number:

Job Reference Number:

For the attention of:

Please find enclosed the above equipment returned to you for the following reason(s):

Before proceeding with any work an estimate of repair is required.



Yours Faithfully

Nurul Amin

**All invoices should be sent to:- NHS Shared Financial Services, Phoenix
House, Topcliffe Lane, Tingley, W Yorkshire, WF3 1WE**