

Repair

total

1 10.9.03 3156
 2 10.9.03 1210 Customer Repair
 5 + 3 Previously repaired in past month

1	8.11.04	Repair (customer)	}	Paid
1	15.11.04	Repair		
1	24.11.04	Repair		
2	9.12.05	Repairs		
1	16.2.05	Repairs		
2	2.3.05	Repairs		
2	16.3.05	" "	}	Paid
1	31.3.05	" "		
1	21.4.05	" "		
4	18.5.05	" "		
2	26.5.05	" "		
3	8.6.05	" "	2	17.8.05
1	15.6.05	" "		?
1	6.7.05	" "	}	<u>Paid</u>
2	20.7.05	" "		
1	27.7.05	" "		
1	3.8.05	" "		(Claimed)

Repairs - No Comment

1	23.8.05		Customer Repair	
2	7.9.05		" "	
31	28.9.05	62	changed Cable	
41	28.9.05	3290	Cable	
1	3.11.05	2640	Calibrate	
1	3.11.05	2392	Cable	
1	3.11.05	3241	Cable	
1	3.11.05	233	Cable	
1	9.11.05	27	Cable	
1	17.11.05	1293	Cable	
1	17.11.05	2931	Display	
1	23.11.05	1118	Water Damage	
1	30.11.05	3004	Recalibrated	
1	30.11.05	3348	Recalibrated	
1	30.11.05	2777	Broken Wire	
1	30.11.05	2889	Faulty Sensor	
1	4.1.06	2878	Drifting changed cable - battery	connector
1	8.3.06	2891	Cable	
1	2-5-6	1519	on/off switch	
1	3-5-6	1885	IC2	
1	17.5.06	817	Switches	
1	28.6.06	2722	Cable, battery Connector	Switch
1	5.7.06		Calibrate	

Paid for

Paid for

Paid 1/4	1	12-7-06	Battery conn Calibration	0024 2863
	1	12-7-06	Cable	2863
	1	27-7-6	Cable	0085
	1	27-7-6	Battery conn	1396
	1	27-7-6	Calibration	1260
	1	27-7-6	Calibration knob	3348
	1	27-7-6	Battery Conn	2257

Paid	1	5-8-06	Cable	2905
	1 1	18-8-06	Cable	817
	1	18-9-06	Cable	1540
	1	23-9-06	Cable	3886
	1	2-11-06	Cable	520
	1	13-11-06	Battery Connector	3332
	1	13-11-06	Battery Connector	3345
	1	13-11-06	Battery Connectors	3338
	1	13-11-06	Cable Battery Connectors	2429
	1	13-11-06	Battery Connectors	3330
	1	13-11-06	Battery Connectors	3279
	1	13-11-06	Battery Connector	3349
	1	26-11-6	Dry Solder Joint	3008
	1	9-1-07	Cable	0567
	15 1	9-1-07	Cable / Battery Conn.	2276

	No	No	Repairs
15.01.07	3606	1	IC 2, Cable
7.02.07	3207	1	IC 2, Butt Connector
20.02.07	1209	1	Cable
20.02.07	3207	1	changed everything.
25.02.07	3694	1	Faulty Cable
25.02.07	2760	1	LCD
27.02.07	3731	1	IC 2 Hong Kong
6.03.07	1653	1	Cable
6.03.07	2375	1	Cable
6.03.07	3505	1	Cable
14.04.07	1186	1	LCD L&H.
26.4.07	3299	1	Cable H
26.4.07	3514	1	Cable H
26.4.07	1574	1	Battery Connector
26.4.07	432	1	serviced & Relabelled H
2-5-7	2902	1	on/off switch
25-6-7	3299	1	Cable Long
25-6-7	3882	1	Cable
25-6-7	1561	1	Cable/on/off switch
25-6-7	12065	1	Cable
25-6-7	700	1	Cable
25-6-7	2372	1	Cable
25-6-7	451	1	LCD
25-6-7	2945	1	LCD
25-6-7	2951	1	LCD
25.6.07	2363	1	Battery Connector
4-7-7	0089	1	Cable
4-7-7	3140	1	recalibrated
6-7-7	3734	1	" "

joined
10/5/07

Helen's Repairs

<u>04.07.07</u>		<u>1367</u>	<u>Cable/batt Connector</u>
	04.07.07	1333	Cable/on-off Switch
	04.07.07	3347	Cable
LS	18.08.07	3754	LCV
LS	18.08.07	3755	Cable
LS	18.08.07	3757	Cable
LS	20.08.07	2022	Battery Cable, on/off Switch
LS	20.08.07	0355	Cable
LS	20.08.07	0947	Cable & battery connector
	08.09.07	3257	Cable
	08.09.07	2609	Cable
	08.09.07	0817	Cable
LS	29.09.07	2891	Cable
	30.10.07	1575	Cable
	30.10.07	1892	Cable
	30.10.07	1819	Cable
	30.10.07	889	Cable
LS	30.10.07	2021	Battery Connector
LS	25.11.07	3018	Cable
LS	6.12.07	0157	calibrated New back wheel
LS	6.12.07	0854	Cable & LCV
LS	8.1.08	0303	on/off Switch + Cable
LS	22.2.08	3141	Calibrate
LS	5.3.08	3280	Broken on/off switch
LS	7.3.08	3514	Cable
LS	15.3.08	1670	on/off Switch
LS	16.4.08	2777	Cable ICZ
LS	16.4.08	3141	Cable ICZ
HL	15.5.08	?	Batt Con
HL	15.5.08	?	Batt Con
HL	15.5.08	?	Batt Con + Cable

IS 29.5.08	2018	Batt Connector + Cable
HL 8.6.08	0825	Cable + Dry Joint
HL 8.6.08	0722	Batt Conn
LS 21.6.08	4031	Failed to Repair {IC1, IC2, IC3, P1, P2, etc
LS 29.7.08	3764	{IC2, IC3, IC2, LC1}
LS 29.7.08	1706	Calibrated Cable
LS 14.8.08	1210	Cable Updated board & LC1
LS 28.8.08	0238	Cable - Redid Wires
LS 4.10.08	1585	LC1 + Cable
LS 17.10.08	0247	Calibrated
LS 17.10.08	2885	Cable
LS 30.10.08	3165	Cable + LC1 Full length Cable
LS 22.11.08	3459	Cable
LS 24.11.08	1876	Cable
LS 24.11.08	2889	Cable
LS 24.11.08	3004	Battery Connector
LS 5.12.08	2371	Calibrated
LS 12.12.08	4066	Cable Back Sticker Adjusted
LS 1.1.09	3345	Checked Cable Connections
LS 1.1.09	3034	Resoldered on/off switch
LS 1.2.09	2490	IC2, Dry Jt. Opposite Tape
LS 1.2.09	3882	Serviced Newables
LS 15.2.09	4066 x2	Cable
LS 15.2.09	2891	Calibration
LS 23.2.09	3768	New Cable + back
LS 23.2.09	4142	IC2. Removed on/off switch
LS 23.2.09	4141	Cable
LS 23.2.09	3736	Calibrated
LS 23.2.09	1885	Cable + On/off Switch
LS 10.3.09	3781	LC3 Wires
LS 15.3.09	1519	LC1 + Cable
		Cable On/off Switch
		Is Unreparable
		Calibrated
		Cable

26.3.09	0199	Cable
2.4.09	3753	? Battery Drain
2.4.09	1832	Battery Cable + LC
2.4.09	1665	Calibrated
17.4.09	0024	Cable 102. 103 ^{Front} _{Stickers}
16.5.09	3995	Calibration

claimed 12.7.09