

Service Repair Sheet SRS67851

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Grye Belza

Company/ Hospital Name

Pilgrim Hospital

Department

Clinical Engineering Department

Position

Admin Support Assistant

Direct Phone

General Phone

01205445205

Opera Account

00000600

Email

Grye.Belza@ulh.nhs.uk

Order Number

Date Received

03/Jun/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 28/May/2021 Kate Griffiths

28/May/2021 Kate Griffiths

Customer would like to return 1 x R22-MED under warranty - "Our technician tried to fit it into V808-Atom Transcapsule and it showed a "sensor fault".

03/Jun/2021 Robert Connor

Received 1 x R-22MED s/n 531997

Ready For quote

Repair Complete Signed

C.Green 3.6.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33742	Oxygen Sensor	0110022	531997	Yes

0140013 x 1 FOC
S/N, SRS, SRN

0110022 x 1 FOC
NEWS/N, SRS, SRN

UPS x 1 FOC