

Service Repair Sheet 67851

Contact Name	Grye Belza
Company/ Hospital Name	Pilgrim Hospital
Department	Clinical Engineering Department
Position	Admin Support Assistant
General Phone	01205445205
Opera Account	00000600
Email	Grye.Belza@ulh.nhs.uk
Date Received	03/Jun/2021
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33742	0110022	531997	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0110022    Oxygen sensor R-22MED

11.2 in air and 56.6 in oxygen. This sensor is nonlinear and replaced under warranty with S/N 601879

---