

# EC Declaration of Conformity

We hereby declare under our sole responsibility  
that the generic product group

## Reusable SMARTsat SpO<sub>2</sub> Sensors

**Sensors for continuous and spot check measurement of functional arterial oxygen saturation (SpO<sub>2</sub>) and pulse rate**

REF	Type	Product
6020132014	0014752	SC7500VM, SoftCap SpO <sub>2</sub> sensor - Large
6020132025	0014754	SCM7500VM, SoftCap SpO <sub>2</sub> sensor - Medium
6020132305	0014753	SCP7500VM, SoftCap SpO <sub>2</sub> sensor - Small
6020132030	0014760	SC7500-LVM, SoftCap SpO <sub>2</sub> sensor - Large, 1.8m cable
6020132012	0014651	SF7500VM, SoftFlap SpO <sub>2</sub> sensor
6020132016	0014851	W7500VM, SoftWrap SpO <sub>2</sub> sensor
6020132264	0014850	EP7500VM, Ear Probe SpO <sub>2</sub> sensor

Complies with the essential requirements of Annex I and Annex II of the Council Directive 93/42/EEC as amended by 2007/47/EC concerning medical devices as well as the requirements of Regulation (EU) 2017/745, Article 120, Chapter (3).

In accordance with Annex IX of the Council Directive 93/42/EEC the product has been classified as Class IIb.

Application of the CE-marking:



MEDCERT GmbH, Pilatuspool 2, 20355 Hamburg

**Validity:**

**Date of expiry: 31 January 2022**

**Issuer:**

bluepoint medical GmbH & Co. KG  
An der Trave 15  
23923 Selmsdorf  
Germany

**Place, Date:**

Selmsdorf, 20 December 2019

**Legally binding signature:**

  
.....  
Bernd Lindner  
General Manager