Service Repair Sheet SRS67842

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Nurul Amin

Bradford Royal Infirmary

Clinical Engineering

Specialist Healthcare Scientist

01274 364147

0127 454 2200

00000620

nurul.amin@bthft.nhs.uk

27/May/2021

Robert Connor

Viamed

Quote

Notes 25/May/2021 Zoey Teal

25/May/2021 Zoey Teal

Sending in a Tom Thumb. Serial number 0401193. The pressure will not go above 10 cmWG. They have replaced the pressure gauge with a new one but made no difference.

27/May/2021 Robert Connor

1 x tom thumb minus gauge s/n 0401193, tom thumb gauge

Ready For quote

Repair Complete Signed

CGNEEN 27-5-2021

CGNOON 7-6:2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33726	Tom Thumb	0310030	0401193	N

0380000 × 1 £90 SIN, SLS, SLN 2.00 0330212 × 1 £150 SLS, SLN

UPS × 1 £10.

Add to order 0330203 × 20 £125 Cg 8-6-2021