

Service Repair Sheet SRS67853

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Lesley McSporran

Dumfries & Galloway Royal Infirmary

Medical Engineering

Medical Engineering Technician Specialist

01387241477

00001480

lesley.mcsporrans@nhs.scot

10/Jun/2021

Robert Connor

Viamed

Quote

Notes 02/Jun/2021 Kate Griffiths

02/Jun/2021 Kate Griffiths

Customer would like to send in Tom Thumb for service.

"It passes all the service tests at 10 l/min. But when set to 5 l/min it does not seem to work correctly. It just had the precision valve replaced. (here by us)." s/n VMCC43

10/Jun/2021 Robert Connor

Received 1 x tom thumb s/n VMCC43

Ready For quote

canon 10-6-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33769	Tom Thumb		VMCC43	2

0380000 x 1 @ £90.

S/N, SRS, SRN

UPS x 1 @ £10.

Select Currency

Pounds | Euros | Usd

Repair SRS67853 Values in Pounds @ 1: 1.00

SRN33769	Tom Thumb	0310030			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		0380000	1	90.00	90.00
				Parts Cost	90.00

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent