

Service Repair Sheet 67866

Contact Name	Ryan Filer
Company/ Hospital Name	Musgrove Park Hospital
Department	Med Electronics
Position	Tech
Direct Phone	0182334 2486
General Phone	01823 333 444
Opera Account	00005110
Email	Ryan.Filer@somersetft.nhs.uk
Date Received	16/Jun/2021
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	Under Warranty
SRN33936	0110425	FK70799060	Oxygen Sensor	

Time :0 Hour(s)

Parts Replaced

Qty: 1    0110425    Oxygen sensor MAX-250. With `O` ring.

The MAX-250 Oxygen sensor has been replaced under warranty with SN:FK70799032.

---