

Service Repair Sheet SRS67865

Contact Name
Company/ Hospital Name
Department
Position
Direct Phone
General Phone
Opera Account
Email
Order Number
Date Received
Booked in By
Main Company
Type Return

Date Promised to Return: 01 Jan 1970 -3600
Alvin Manaloto
Eastbourne District General Hospital
EME Department
EME Technician
03001315151 Ext: 735151

00004621
alvin.manaloto@nhs.net

21/Jun/2021
Robert Connor
Viamed
Warranty

Notes 11/Jun/2021 Kate Griffiths

11/Jun/2021 Kate Griffiths

Customer would like to send AX300, s/n 334659. It "always bleeps and is not giving correct reading". Warranty.
(Purchased by Meadowfield Hosp)

21/Jun/2021 Robert Connor

1 x AX300 s/n 334659, with base stand and 3 x AA battery, cable s/n 0012426, R17MED s/n 593247, with flow diverter
One of the battery pins is bent so one battery is loose in the compartment.

Ready For quote

Repair Complete Signed

CGreen 23-6-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33939	Oxygen Monitor	0111230	334659	Yes
SRN33940	Oxygen Monitor Cable	0131202	0012426	Yes
SRN33941	Oxygen Sensor	0110017	593247	Yes

+ sold 6-12-19
- sold 6-12-19
- sold 30-3-2021

0140006 x 1 FOC
S/N, SRS, SEN
NO fault found.

0140025 x 1 FOC
S/N, SRS, SEN - NO fault found

0140013 x 1 FOC
S/N, SRS, SEN.

→ 0110017 x 1 FOC
NEW S/N, SRS, SEN

UPS x 1 FOC

EQUIPMENT RETURN FORM

TO: VIAMED Repairs Team SRS67865 15 Station Road, Cross Hills, Keighley, West Yorkshire, BD20 7DT		DATE: 16.06.2021
DESCRIPTION of equipment: AX300, s/n 334659		
Contamination details: Has been exposed to blood, body fluid etc <input type="checkbox"/> No Has been disinfected/cleaned <input type="checkbox"/> Yes Method of Disinfection/cleaning used: CLINNEL WIPES GREEN TUB		
DATE of purchase (approx.): 6/12/2019		
REASON FOR RETURN: It was reported by clinician that always bleeps and not giving correct readings.	ACTION REQUIRED: REPAIR/SERVICE UNDER WARRANTY ISSUE.	
OUR Ref: JOB NO: 604900 EME ASSET NO: 101126		
CONTACT: ALVIN MANALOTO alvin.manaloto@nhs.net EME TECHNICIAN EASTBOURNE DISTRICT GENERAL HOSPITAL EAST SUSSEX HEALTHCARE NHS TRUST KINGS DRIVE EASTBOURNE EAST SUSSEX BN21 2UD		
TEL: 0300 131 5151	FAX:	e-mail:

Any Person Identifiable Information (PII) that this device may hold is the property of East Sussex Healthcare Trust (ESHT are the Data Controller for the PII on this device). Any access or processing of any PII on this device by you, your company, its servants or agents may breach Data Protection legislation. By being in receipt of this device you are bound by ESHT's confidentiality processes.

Further information regarding PII is available from Esh-tr.infogov@nhs.net