

Service Repair Sheet SRS67758

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Jim McKay

JFD Limited

Senior QC Inspector

01224 740 145

VST036

06/Apr/2021

Robert Connor

VandagraphST

Warranty

0140010 x 53
S/N, SRS, SRN
Returned as instructed
by Duncan Clark.

Notes 08/Mar/2021 Derek Lamb

06/Apr/2021 Robert Connor

Cells arrived 2021-04-06 repurposed unused blank SRS.

Received 58 x JFD sensor, failed magnetic range testing.

06/Apr/2021 Robert Connor

Cells added to RMA box 774

07/Apr/2021 Catherine Spence

with steve nixon

22/Apr/2021 Catherine Spence

on hold with SN.

We have sent back 5 sensors s/n:101906,109607,101949,102093,102148 back to supplier as per steve nixon ro be

looked at returned 14/04/2021 on RMA300237032

10/May/2021 Catherine Spence

five units out with supplier

19/May/2021 Catherine Spence

with supplier

01/Jun/2021 Catherine Spence

five items still out with supplier

10/Jun/2021 Catherine Spence

with supplier

23/Jun/2021 Catherine Spence

with supplier

23/Jun/2021 Catherine Spence

asked Zoey to chase customer again

23/Jun/2021 Steve Nixon

Duncan Clark reminded again and he has instructed for us to return the sensors. From the above it looks like we can only return 53 out of the 58 sensors. Five have been dismantled for assessment.

Failing mag range testing is outside of the warranty provisions.

0140010 x 5
S/N: 101949 - SRN 33556
S/N: 101906 - SRN 33564
S/N: 101907 - SRN 33565
S/N: 102093 - SRN 33606
S/N: 102148 - SRN 33608

Dismantled as instructed
by Duncan Clark.

UPS x 1 FOC

Ready For quote

Repair Complete Signed

C. Neen 1-7-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33551	Oxygen Sensor VST	8010050	101940	
SRN33552	Oxygen Sensor VST	8010050	101941	
SRN33553	Oxygen Sensor VST	8010050	101942	

See
other
sheets.

Cathy can you please return the following to JFD as instructed by Duncan Clark.: SRS67758 58 sensors - Failed JFD mag range testing which is outside of the scope of warranty. 5 have been dismantled, so 53 returned. SRS67765 3 sensors - Corrosion - out of warranty.

Items for Repair

SRN	Equipment	Stock Reference	Serialnumber Barcode	Serialnumber	Description	Warranty New	Warranty Repair	Status	Repair Code	Time (H:M)
SRN33551	Oxygen Sensor VST	8010050	1649314	101940	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33552	Oxygen Sensor VST	8010050	1649315	101941	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33553	Oxygen Sensor VST	8010050	1649316	101942	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33554	Oxygen Sensor VST	8010050	1649321	101947	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33555	Oxygen Sensor VST	8010050	1649322	101948	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33556	Oxygen Sensor VST	8010050	1649323	101949	Oxygen Sensor - JFD	No	No	End Of Life Out to Supplier	High magnetic signature	00:00
SRN33557	Oxygen Sensor VST	8010050	1649360	101986	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33558	Oxygen Sensor VST	8010050	1649361	101987	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33559	Oxygen Sensor VST	8010050	1649291	101917	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33560	Oxygen Sensor VST	8010050	1649367	101993	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33561	Oxygen Sensor VST	8010050	1649368	101994	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33562	Oxygen Sensor VST	8010050	1649369	101995	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33563	Oxygen Sensor VST	8010050	1649353	101979	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33564	Oxygen Sensor VST	8010050	1649280	101906	Oxygen Sensor - JFD	No	No	End Of Life Out to Supplier	High magnetic signature	00:00
SRN33565	Oxygen Sensor VST	8010050	1649281	101907	Oxygen Sensor - JFD	No	No	End Of Life Out to Supplier	High magnetic signature	00:00
SRN33566	Oxygen Sensor VST	8010050	1649328	101954	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33567	Oxygen Sensor VST	8010050	1649329	101955	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33568	Oxygen Sensor VST	8010050	1649302	101928	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00

SRN33569	Oxygen Sensor VST	8010050	1649303	101929	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33570	Oxygen Sensor VST	8010050	1649337	101963	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33571	Oxygen Sensor VST	8010050	1649288	101914	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33572	Oxygen Sensor VST	8010050	1649289	101915	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33573	Oxygen Sensor VST	8010050	1649284	101910	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33574	Oxygen Sensor VST	8010050	1649285	101911	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33575	Oxygen Sensor VST	8010050	1649342	101968	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33576	Oxygen Sensor VST	8010050	1649343	101969	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33577	Oxygen Sensor VST	8010050	1649308	101934	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33578	Oxygen Sensor VST	8010050	1649309	101935	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33579	Oxygen Sensor VST	8010050	1649310	101936	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33580	Oxygen Sensor VST	8010050	1649283	101909	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33581	Oxygen Sensor VST	8010050	1649293	101919	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33582	Oxygen Sensor VST	8010050	1649334	101960	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33583	Oxygen Sensor VST	8010050	1649349	101975	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33584	Oxygen Sensor VST	8010050	1649347	101973	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33585	Oxygen Sensor VST	8010050	1649344	101970	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33586	Oxygen Sensor VST	8010050	1649365	101991	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33587	Oxygen Sensor VST	8010050	1649306	101932	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33588	Oxygen Sensor VST	8010050	1649295	101921	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33589	Oxygen Sensor VST	8010050	1649355	101981	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33590	Oxygen Sensor VST	8010050	1649313	101939	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00

SRN33591	Oxygen Sensor VST	8010050	1649297	101923	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33592	Oxygen Sensor VST	8010050	1649363	101989	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33593	Oxygen Sensor VST	8010050	1649340	101966	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33594	Oxygen Sensor VST	8010050	1649332	101958	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33595	Oxygen Sensor VST	8010050	1649375	102001	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33596	Oxygen Sensor VST	8010050	1649379	102005	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33597	Oxygen Sensor VST	8010050	1649383	102009	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33598	Oxygen Sensor VST	8010050	1649384	102010	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33599	Oxygen Sensor VST	8010050	1649393	102019	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33600	Oxygen Sensor VST	8010050	1649395	102021	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33601	Oxygen Sensor VST	8010050	1649396	102022	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33602	Oxygen Sensor VST	8010050	1649398	102024	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33603	Oxygen Sensor VST	8010050	1649399	102025	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33604	Oxygen Sensor VST	8010050	1681767	102071	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33605	Oxygen Sensor VST	8010050	1681760	102064	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33606	Oxygen Sensor VST	8010050	1681789	102093	Oxygen Sensor - JFD	No	No	End Of Life Out to Supplier	High magnetic signature	00:00
SRN33607	Oxygen Sensor VST	8010050	1681762	102066	Oxygen Sensor - JFD	No	No	Return Unrepaired	High magnetic signature	00:00
SRN33608	Oxygen Sensor VST	8010050	1681844	102148	Oxygen Sensor - JFD	No	No	End Of Life Out to Supplier	High magnetic signature	00:00

Currently Viewing ITEM SRN33551 Oxygen Sensor VST 8010050 102148 Oxygen Sensor - JFD

[Create a Meeting Issue](#)

Items Requested for Repair	Item notes	Engineer Notes	Notes To Customer
	Sensor reported as failed mag range testing, this is outside of the scope of warranty. Sensor dismantled, unable to	23/Jun/2021 Steve Nixon New Warranty Flagged OFF Sensor reported as failed mag	Sensor reported as failed mag range testing, this is outside of the scope of warranty. Sensor dismantled, unable to

