

Invoice number: RST130628-1
Shipment date: 28/05/2021

Magnetic signature analysis of production components

1000 x 8030450 DB200381 Oxygen Sensor Cathode

Once the analysis has been carried out, please complete the following and email back to VST:

info@vandagraphst.com

Magnetic signature values:

Highest: 0.21 nT

Lowest: 0.19 nT

Result (PASS/~~FAIL~~):

Name: Ian Honeyman
Position: Magnetic Ranger

Signed by:
On behalf of JFD *Ian Honeyman*

Date: 24/06/2021

Vandagraph Sensor Technologies Limited
15 Station Road, Cross Hills, Keighley, West Yorkshire BD20 7DT, United Kingdom
Tel +44 (0) 1535 634900 Fax +44 (0) 1535 635582 e-mail: info@vandagraphst.com
www.vandagraphst.com

Company registered in England, No. 03116432