12/07/2021 Gmail - New P.I



Sarah Walton <viamed.sarah.walton@gmail.com>

New P.I

1 message

bmg.equipment
 bmg.equipment@gmail.com>

11 July 2021 at 16:24

To: sarah.walton@viamed.co.uk

Dear Sarah,

Good morning!

Please send us P.I for:

nr.	Part number	Description	Quantity
1	0110429	O2 SENSOR MAX O2 250E	5
2	0110023	R-23V Oxygen Sensor , Draeger Evita 2	10

B.R

Anastasios

Στις 6/7/2021 1:24 μ.μ., ο/η Sarah Walton έγραψε:

Dear Anastasios,

Good morning!

Thank you for your email.

Apologies. It looks like they have made recent changes to the design of their sensor making the input thread slightly larger, so unfortunately our sensor is no longer compatible.

Please could you return this sensor back to Viamed and we will issue you with a credit once we have received it?

Kind regards

Sarah Walton

Customer Service Representative

Please note: We are currently working remotely however telephone calls to the main office will still be answered.

Our office opening hours are Monday - Friday 8.30am - 5.30pm.

http://www.viamed.co.uk

Email sarah.walton@viamed.co.uk

GDPR / Privacy Policy available at www.viamed.co.uk/policy

Tel: +44 (0)1535 634542 Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up

12/07/2021 Gmail - New P.I

Twitter: twitter.com/ViamedLtd

Facebook Page: facebook.com/ViamedLtdUK

Company registered in England, No. 01291765.

E-mail Disclaimer

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Tue, 6 Jul 2021 at 07:55, bmg.equipment

bmg.equipment@gmail.com> wrote:

Dear Sarah, Ryan

Good morning!



We have a problem with o2 sensor for SIARE.

On your webpage it is suggested that o2 sensor R-17V is compatible with Siare equipment.

we sent our client a R-17V but he informed us that it does not "fit" in the machine(see att.)

Please commnet

Thanks in advance

B.R

Anastasios