

Service Repair Sheet SRS67279

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Thibaut Bister

Minerve

+33 326425015

00006280

t.bister@minerve.eu

29/Oct/2019

Robert Connor

Viamed

Quote

Notes 17/Oct/2019 Ryan Swaine

17/Oct/2019 Ryan Swaine

17/Oct/2019 Ryan Swaine

VM-2500-M VET (SN: A1711150006)

Not recognising the airway adapters, I suspect that it has been continually used with an old adapter and had gone of the calibration scale.

I would suggest we try to bring it back in to spec with some discoloured airway adapters.

29/Oct/2019 Robert Connor

Vm-2500-M-Vet s/n A1711150006 with full case and accessories including IRMA s/n 247009

Ready For quote

CGN00N 24-1-2020

Repair Complete Signed

CGN00N 12-7-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN32016	Capnograph	4410530	A1711150006	Yes No
SRN32017	Accessory	4420505	247009	Yes No

NO? was in warranty when received eg 10th Feb 2020

4440000 x 1 @ €0.00
S/N, SRS, SRN

4440010 x 1 @ €90.00 - SRN 32017
S/N, SRS, SRN

4420505 x 1 @ €950.00
SRS, SRN

UPS6 x 1 @ €10.00
61x47x25cm

Insurance €9.50

