

Service Repair Sheet SRS67888

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Janice Hunter

Company/ Hospital Name

Hunter and Co

Department

Accounts

Position

Direct Phone

01383 852049

General Phone

01383 852049

Opera Account

00020062

Email

hunterandco@outlook.com

Order Number

Date Received

13/Jul/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 09/Jul/2021 Steve Hardaker

09/Jul/2021 Steve Hardaker

09/Jul/2021 Steve Hardaker

Returning some R-22A sensors that are reading high. No serial numbers known as yet but she thinks they are under warranty.

12/Jul/2021 Steve Hardaker

09/Jul/2021 Steve Hardaker

09/Jul/2021 Steve Hardaker

09/Jul/2021 Steve Hardaker

Returning some R-22A sensors that are reading high under warranty.

s/n 544445, 544446 - sold 19/11/20

s/n 573327 - sold 1/3/21

13/Jul/2021 Robert Connor

Received 3 x R22A s/n 544446, 544451, 573327

Ready For quote

C. Green 19.7.2021

Repair Complete Signed

C. Green 19.7.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN33996	Oxygen Sensor	0110122	544446	Yes
SRN33997	Oxygen Sensor	0110122	544451	Yes
SRN33998	Oxygen Sensor	0110122	573327	Yes

0140013 x 3 FOC  
old S/N, SR5-

0110122 x 3 FOC  
new S/N, SR5

UPS x 1 FOC