

Service Repair Sheet SRS67887

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Manuela Kellner

Company/ Hospital Name

Hellmut Habel Gesellschaft m.b.H.

Department

Purchase Dept

Position

Direct Phone

+43/1/292 66 42 28

General Phone

Opera Account

Email

manuela.kellner@habel-medizintechnik.at

Order Number

Date Received

15/Jul/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 09/Jul/2021 Sarah Walton

09/Jul/2021 Sarah Walton

serial number 571933

Failure: no O2 measurement, 0 Volt output voltage

Relates to RVM128858-1

15/Jul/2021 Robert Connor

Received 1 x R17MED s/n 571933, with flow diverter.

Ready For quote

Repair Complete Signed

CGNEON 19.7.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34005	Oxygen Sensor	0110017	571933	Yes

0140013 x 1 FOC
old S/N, S/N, SRS.

0110017 x 1 FOC
new S/N, S/N, SRS.

UPS x 1 23x15x12cm 0.20kg

Ship now or with
next order? eg 19.7.2021

VIAMED
RMA# SRS67887

15 Station Road, Cross Hills, Keighley
BD20 7DT WEST YORKSHIRE
ENGLAND

Hellmut Habel Gesellschaft m.b.H.
Ignaz-Koeck-Strasse 20
1210 VIENNA
AUSTRIA / EUROPE
Tel.: + 43 / 1 / 292 66 42 - 0
Fax: + 43 / 1 / 290 12 45
einkauf@habel-medizintechnik.at

Enquiry no. L078285

Date: 09.07.2021

Your phone no.	+44 1535 634542
Your fax no.	+44 1535 635582
Your supplier no.	301782 GB287389593

DELIVERY DATE

Contact:	Manuela Kellner
Our Customer No.:	I0007025
Page:	1 / 1

Delivery terms: ex works

Shipment: UPS Standard

We herewith return due to SRS67887 for investigation.

Part No.	Description	Qty.	Unit	Price	%	Total
99REKLAMATION	Viamed #0110017; ser.nr. 571933, SRS67887 (99REKLAMATION our ref.) Failure: no O2 measurement, 0 Volt output voltage	1,0		0,00		0,00

"Steuerfreie innergemeinschaftliche Lieferung
gem. Art.6, Abs.1, iVm Art 7 UStG"

*** Proforma Rechnung:L078285***

Order Value	Others		Base	Tax %	Tax Value	Total
0,00	0,00		0,00	0,00		0,00
Payment: net 30					EUR	0,00

www.habel-medizintechnik.at