

Invoice Correction 5256957605**ORIGINAL**

Invoice Correction Date: 29.07.2021

Purchase Order No: RTS3100803

EnviteC-Wismar GmbH
 Alter Holzhafen 18
 23966 Wismar
 Germany

Bill to: 1797448
 Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

Sold to: 1797448
 Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

Order Information:

Credit Memo Request: 63644348
 Sales Office: 2170 DE SPS SIOT EnviteC
 Incoterms: FCA Origin
 Payment Terms: Net 30 Days After Invoice Date

Ship to: 1797448

Vandagraph Sensor Technologies Ltd.
 15 Station Road
 Cross Hills, West Yorkshire
 KEIGHLEY
 BD20 7DT
 United Kingdom

ITEM NO.	Material No.	QTY	UOM	UNIT PRICE	EXT. PRICE	VAT %
000010	E1002461 Oxygen Sensor OOD103-JFD Customer Part Nbr: 8010050 credit note for SN: 102258	1	EA	77,78	-77,78	0%

Totals

Total Before VAT -77,78 EUR
 VAT-Total 0,00 EUR
 Total Amount -77,78 EUR

Tax Information

VAT Reg. No: GB991168000
 Ship To: 0001797448
 Tax Exempt Reason: Tax free export - Goods

Integrity and Compliance at Honeywell

Honeywell places the highest value on the integrity of its business operations. We expect our employees and business partners to comply with our Code of Business Conduct and all applicable laws and regulations. If you become aware of illegal or improper conduct by a Honeywell employee please contact your Honeywell relationship manager. You can also report misconduct directly to us either via e-mail to our Integrity and Compliance Office at access.integrity.helpline@honeywell.com or via our Access helpline accessible worldwide at 800-237-5982.

Duns #:

Bank Details: No Payment Due

Account Name:

Account No:

Bank Key:

Swift Code:

BAN:

Payment method:

Registered No: Schwerin HRB 1838

VAT Reg No: DE137437213

Federal ID #:

Registered Address: EnviteC-Wismar GmbH, Alter Holzhafen 18, Wismar, 23966, Germany

Credit Analyst: Claudia Voinea + 40 316302631

Sales Rep: Jessica Wagner + 49 3841 360 206

Customer Care Rep: Jessica Wagner + 49 3841 360 206

EnviteC-Wismar GmbH
 a Honeywell Company

WEEE Reg.No.: 57798680
 Managing Director: Marcus Ostländer

Internet: www.envitec.com
 email: info@envitec.com

Delivery exclusive to our General Terms (February 2016). We retain title to our goods and deliveries until all the payment from the sales contract have been received. Your order will be executed exclusively to our terms and conditions. These can be sent by separate request or viewed at www.envitec.com

PRÜFBERICHT - TEST REPORT

Sauerstoffsensoren MED / Oxygen Sensors Medical Use

Vorgangsnummer/ Process Number	47610351	Benachrichtigungsnummer/ Notification Number	300240868
-----------------------------------	----------	---	-----------



Kunde / Customer <i>Vanda graph</i>	Kunden - Auftragsnummer / Customer Order No. RTS3100803
--	---

Auftragsdatum / Date of Order	22.07.2021
-------------------------------	------------

Lief.an code/ Delivery Code	Aufgenommen von / Received by <i>KG</i>
--------------------------------	--

Beanstandung / Objection

<i>Signal zu hoch</i>

Artikelnummer / Part No.	Herstell. Datum / Date of Manufact.	Code + Seriennummer / Code + Serial No.	Garantie bis / End of Warranty Date	Beschreibung / Description
E1002461	<i>02121</i>	102258		Oxygen Sensor OOD103-JFD

Prüfkriterien unter Standardbedingungen / Test Parameters under Standard Conditions

- | | | |
|---|---|------------------|
| 1 | Ausgangssignal an Luft / Output Signal in Atmosphere | [20,95 % vol.O2] |
| 2 | Linearitätsfehler, Meßfehler / Linearity Error, Error of Measurment | [100 % vol.O2] |
| 3 | Nullsignal / Zero Signal | [100 % vol.N2] |

Messprotokoll / Test Report			
Prüfkriterien / Test Parameters	Spezifiziert / Specified	Gemessen / Measured	
		Signal1	Signal1
1 [mV]	<i>6.5-9</i>	<i>43</i>	
2 [%]	<i>6200</i>		
3 [µV]	<i>63</i>		

Prüfbericht / Checklist		
Fehler / Defect:	O2-S001..Luftwert außerh.Toleranz / Signal Out Of Tolerance	☒
	O2-S002..Signal Instabil / Signal Not Stable	☐
	O2-S003..Offset Außerh. Toleranz / Offset Out Of Tolerance	☐
	O2-S004..Signal 0mV / Signal 0mV	☐
	O2-S005..Gehäuse Defekt / Housing Defective	☐
	O2-S006..Elektrolytaustritt / Electrolyte Leakage	☐
	O2-S007..Sonstiges / Others	☐
Bemerkungen / Remark	kein Fehler festgestellt, Sensor innerhalb Spezifikation / No failure found, Sensor within specifications	☐
	außerhalb der Garantie / out of warranty	☐

Anerkannt / Accepted ☒	Nicht anerkannt / Not Accepted ☐
--	---

Austausch / Replacement	Rücksendung / Return
VB-Nr. / VB No	Lieferdatum / Delivery Date
Seriennummer / Serial No.:	Geprüft / Inspected: <i>Bue</i>
Lieferdatum / Delivery Date Bearbeitet / Processed <i>30.07.21</i>	