

EC Declaration of Conformity

We hereby declare under sole responsibility that the product

VM 2500-M and VM 2500-S

hand held capnograph and pulse oximeter for continuous and spot-check monitoring of end-tidal CO₂ concentration (EtCO₂), inspired CO₂ concentration (FiCO₂), functional arterial oxygen saturation (SpO₂) respiration rate and pulse rate,

Product No.

4410500, 4410501, 4410520 and 4410521

conform with the essential requirements of Annex II of the Council Directive 93/42/EEC of 14 June 1993 concerning medical devices.

In accordance with Annex IX of the Directive 93/42 EEC the product has been classified as Class IIb

Application of the CE-marking:

CE 0086

Issuer:

Viamed Ltd.
15 Station Road
Cross Hills
Keighley
West Yorkshire, BD20 7DT
United Kingdom

Place, Date:

Keighley, 4 August 2009

Legally binding signature:

Derek Lamb (Managing Director)