

Calibration Certificate

The Maxtec MAXO2ME Oxygen analyser has been tested for accuracy by using a UKAS certified voltage source.

Calibrated voltage source	Microcal
ISO calibration index number	CE076
Serial number of the Microcal	6939K6
Certificate number of the Microcal	1646861
Serial number of the MAXO2ME	FA75699022
Certificate number of the MAXO2ME	1629716938

Voltage input:	Target reading:	MAXO2ME reading:
4.78mV	10%	9.9%
9.99mV	20.9%	20.8%
10.0mV	21%	20.8%
11.95mV	25%	25%
23.9mV	50%	50%
35.85	75%	75%
47.8mV	100%	100%

Tested by: M.Green

Signed: