Service Repair Sheet SRS67933

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Alan Marples

Chesterfield Royal Hospital

Department of Clinical Engineering

Specialist Clinical Engineer

01246 513065

01246 277 271

00001170

alanmarples@nhs.net

23/Aug/2021

Robert Connor

Viamed

Warranty

Notes 19/Aug/2021 Zoey Teal 19/Aug/2021 Zoey Teal

Returning a R-23V that is faulty. Serial number V121721.

23/Aug/2021 Robert Connor

Received 1 x R-23V s/n V121721, looks to be a large bubble in the electrolyte.

Ready For quote

Repair Complete Signed

Careen 24-8-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34090	Oxygen Sensor	0110023	V121721	Yes

0140013 × 1 FOC 010 5/N, 5RS, SRN

0110023 × 1 FOC New S/N, SRS, SRN

UPS × 1 FOC.