

Viamed Limited · 15 Station Road · Cross Hills
Keighley · West Yorkshire · BD20 7DT · United Kingdom
Tel: +44 (0)1535 634542/636757 Fax: +44 (0)1535 635582
Email: info@viamed.co.uk Website: www.viamed.co.uk

Service Repair Sheet 67920

Contact Name

Company/ Hospital Name

Direct Phone General Phone

Opera Account Email

Date Received Booked in By Lisa Vickers

Medivent Ltd (Euro)

0162 80338

00353 16280338

00007374

info@medivent.ie

12/Aug/2021

Robert Connor

Repair

SRN34073

Ref

S/N

4410520 A1809150030

Equipment Type

Capnograph

Under Warranty

Complete - Repaired

Time: 0 Hour(s)

The Capnograph VM 2500-S has been repaired. The OLED connector has been fixed in place.