

Service Repair Sheet SRS67930

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Magnus Persson

Company/ Hospital Name

Swevet - AB

Department

Position

Direct Phone

46 416 258 21

General Phone

46 771215500

Opera Account

00006308

Email

magnus.persson@swevet.se

Order Number

Date Received

01/Sep/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 17/Aug/2021 Zoey Teal

17/Aug/2021 Zoey Teal

Sending in 2 x 0111263 which is are showing error code E06. Serial numbers FC07299007 and FC07299009

01/Sep/2021 Robert Connor

Received 2 x MAXO2ME s/n FC07299007, with cable and sensor s/n FC24599068, FC07299009, with cable and sensor s/n FC24599058

Ready For quote

CGreen 1-9-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34104	Oxygen Monitor	0111263	FC07299009	2
SRN34103	Oxygen Monitor	0111263	FC07299007	2
SRN34105	Oxygen Sensor	0110452	FC24599068	2
SRN34106	Oxygen Sensor	0110452	FC24599058	2

DIST

Ins ?

0131700 x 2 - SRN: 24.23 € each
SLS, SRN

9950054 x 1 2.24 €
SLS, SRN

UPS x 1 rechargeable 22x24x12cm about 1kg



Proforma invoice No F8006

Ship to: Date 2021-08-17

Shippers name:

Viamed Ltd
15 Station Road
Cross Hill Keighley
BD20 7DT West Yorkshire
Contact : Zoey Teal

Swevet AB
Forskaregatan 1D

SE-275 37 SJÖBO
Contact: Magnus Persson
Mail: magnus.persson@swevet.se

<http://www.viamed.co.uk>
Email: zoey.teal@viamed.co.uk
Tel: +44 (0)1535 634542
Fax: +44 (0)1535 635582

This is veterinary equipment for repair, Equipment will return.

Nr	Description	
4410540	Mextec Oxygen measure equipment	2 Pc

Value : 901 EUR

Magnus Persson
2021-08-17

Tel +46 416-25821

Orgnr 556621-5843
VAT SE556621584301
magnus.persson@swevet.se

Tillverk land : GB
Vikt : 3kg
Tariffnr : 90189084
Eorin nr: SE5566215843

**Select Currency**

Pounds | Euros | Usd 34105 Complete - Repaired
34106 Complete - Repaired

Quote 1 being Generated : not complete notes below

Repair SRS67930 Values in Euros @ €1.17470: ♦1.00

SRN34104	Oxygen Monitor	0111263			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		0131700	1	€35.24	€35.24
		9950054	1	€2.58	€2.58
SRN34103	Oxygen Monitor	0111263			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		0131700	1	€35.24	€35.24
				Parts Cost	€73.07

DIST
AS per Instruct.

23.23 € cg 1-9-2021
2.24 € 1-9-2021

24.23 € cg 1-9-2021

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent