

Service Repair Sheet SRS67946

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Zainab Jaffer

Barnet Hospital MES Dept

Equipment Librarian

0208 2164978

00000261

zainab.jeffer.ext@siemens-healthineers.com

08/Sep/2021

Robert Connor

Viamed

Quote

Notes 08/Sep/2021 Robert Connor

08/Sep/2021 Robert Connor

1 x V1000 s/n PR01866A10, battery contacts corroded from water ingress

Ready For quote

CGreen 9-9-2021

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34185	Foetal Simulator	1410000	PR01866A10	N

1480000 x 1 £45
S/N, SRS, SRN.

1430309 x 1 FOC
SRS, SRN

1430288 x 1 £2.50
SRS, SRN

UPS x 1 £10

EQUIPMENT RETURNS PENDING NUMBER :

2021/200

RMA / REF:

SHIP TO
VIAMED LTD
15 STATION ROAD, CROSSHILLS
KEIGHLEY
WEST YORKSHIRE
BD20 7DT
0153 563 4542

JOB NUMBER: 433981

ASSET No: 1010009149

DESCRIPTION: Foetal Heart Simulator

SERIAL No: PR01866A10

MODEL: V1000

FAULT DESCRIPTION	Water ingress
--------------------------	---------------

ACCESORIES	N/A
-------------------	-----

Please advice cost prior to carrying out repairs.

Please note that a service report should accompany the equipment where appropriate.
To avoid delay in payment the invoice should be sent to Siemens plc c/o Data Capture Solution,
PO Box 4728, Slough, SL1 0AE.

Yours Faithfully,

Zainab Jaffer
Equipment Librarian
zainab.jaffer.ext@siemens-healthineers.com
Please Cc: Barnet.MES.med.gb@siemens.com
Tel: 020 8216 4978



Select Currency

Pounds | Euros | Usd

Repair SRS67946 Values in Pounds @ £1: £1.00

SRN34185	Foetal Simulator	1410000			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		1430288	1	£2.50	£2.50
		1480000	1	£45.00	£45.00
		1430309	1	£15.00	£15.00
				Parts Cost	£62.50

FOC cg 9.9.21

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent