

### **Classification Rationale**

<b>Product Family</b>	<b>“TEE” ADAPTERS</b>		
<b>Type</b>	<b>Rule</b>	<b>Class</b>	<b>Additional Information</b>
<b>Non-Invasive</b>	Yes – Rule 9	<b>IIa</b>	Allows exchange of gases between patient an monitor
<b>Active</b>	Yes		
<b>Invasive</b>			
<b>Special</b>			

BSI Flow chart includes: **Yes** or No

Classified by: .....

Date: ...../...../.....