

## EC Declaration of Conformity

We hereby declare under sole responsibility that the product

**Reusable Temperature Probe,**

Product No.

**0212009, 0212010, 0212011, 0212012, 0212013, 0212014, 0212015, 0212016,  
0212017, 0212018, 0212019, 0212020, 0212021, 0212022, 0212023, 0212024 ,  
0212025, 0212026, 0212027, 0212028, 0212029, 0212030, 0212031, 0212032**

conforms with the essential requirements of Annex II of the Council Directive 93/42/EEC concerning medical devices considering the amendments by the directive 2007/47/EC.

In accordance with Annex IX of the Directive 93/42 EEC considering the amendments by the directive 2007/47/EC the product has been classified as Class IIa

Application of the CE-marking:

CE 0086

**Issuer:**

**Viamed Ltd.**  
15 Station Road  
Cross Hills  
Keighley  
West Yorkshire, BD20 7DT  
United Kingdom

**Place, Date:**

Keighley, 16 May 2010

**Legally binding signature:**

.....

---

Derek Lamb (Managing Director)