

Service Repair Sheet SRS67960

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Philip Tan

Company/ Hospital Name

Junnimed

Department

Position

Direct Phone

+65 65467737

General Phone

Opera Account

00007983

Email

philip_tan@junnimed.com

Order Number

Date Received

30/Sep/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 24/Sep/2021 Sarah Walton

24/Sep/2021 Sarah Walton

SN: 605930

This piece has got unstable O2 reading which can fluctuate. Sometimes reading can go down to 18%.

SN: 617208

This piece was not able to do calibration. After completed count down to zero, the display shows 0.00.

The above are notes from the customer

30/Sep/2021 Robert Connor

Received 2 x R17MED s/n 605930, 617208

Ready For quote

Repair Complete Signed

CGNEEN 5-10-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34212	Oxygen Sensor	0110017	605930	Yes
SRN34213	Oxygen Sensor	0110017	617208	Yes

0140013 * 2 @ FOC
Old SIN: SRS, SRN.

0110017 * 2 @ FOC
New SIN, SRS, SRN

UPS * 1 FOC

Ins?

check with
customer to
see if they
want now or
with next
order.

CG 5-10-2021



JUNNIMED SERVICES PRIVATE LIMITED

MAIN : 15 Changi North Street 1, #01-20 I-Lofts @ Changi, Singapore 498765

SHOP : 62 Bukit Timah Road, #B1-02, Singapore 229902
Little India MRT Station - Downtown Line (DT12) *Along Exit D & E*

TEL. : MAIN +65 6546 7737 • SHOP +65 6514 1243 | FAX : +65 6546 7747

EMAIL : enquiries@junnimed.com • shop@junnimed.com | WEB : junnimed.com

R.O.C : 200403128H

G.S.T: 200403128H

PROFORMA INVOICE

SENDER

Philip Tan/ Lilian Lee
JUNNIMED SERVICES PTE LTD
15 CHANGI NORTH STREET 1
#01-20
I-LOFTS @ CHANGI
SINGAPORE 498765

RECIPIENT

Sarah Walton
Viamed Limited
15 Station Road - Cross Hills
Keighley - West Yorkshire
BD20 7DT, UK

DATE: 27 September 2021

<u>Quantity</u>	<u>Description</u>	<u>Unit Price (USD)</u>	<u>Total Price (USD)</u>
2 units	Faulty Sensor (SN No:617208 & 605930)	\$5.00	\$10.00

Return Reference Number: SRS67960.

COO: UK

Shipment of no commercial value
Value for custom purposes only

Authorised Signature