



Main Account <viamedinbox@gmail.com>

Re: New order 06OCT2021

Mr. Andras Arnold STEINER / MEDICOR Hungary <a.steiner@medicor.hu>
To: zoey.teal@viamed.co.uk, Jutas Deák <djutas@medicor.hu>

7 October 2021 at 16:44

Dear Zoey,

Thank you for your quick response.

Please, go on with the order as we need this shipment a.s.a.p.

Changing the oxygen sensor will take some development and time. Jutas will contact you in this matter.

Best regards

Andras



2021.10.07. 16:55 keltezéssel, Zoey Teal írta:

Dear Jutas,

Thank you for your email.

I have spoken with my Sales Manager and he has confirmed that the medical alternative for the KE-25 is the MAX-25 sensor (part number 0110466).

Both of these sensors have the same specification of 10-15.5mv. I have attached both spec sheets below for your comparison.

I will place your order on hold until you advise me how you would like to proceed.

If you have any further queries, please do not hesitate to contact me.

Kind regards

Zoey Teal

Please note: My working days are Monday to Thursday, 7.45am-6.00pm.

We are currently working remotely however telephone calls to the main office will still be answered.

<http://www.viamed.co.uk>

Email: zoey.teal@viamed.co.uk

Alternate email: viamed.zoey.teal@gmail.com

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up

Twitter: twitter.com/ViamedLtd

Facebook Page: facebook.com/ViamedLtdUK

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.
It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

On Thu, Oct 7, 2021 at 3:30 PM Jutas Deák <djutas@medicor.hu> wrote:

Dear Zoey Teal,

I'm writing on behalf of Mr Andras Steiner.

Thank you for your information!

Please help us to find the best medically acceptable replacement of the KE-25 oxygen sensor.

Which has a similar analog output (11- 13 mV at 21% O2), and the diameter is more or less the same (23 mm).

Best regards,

Jutas Deak

Mr. Andras Arnold STEINER / MEDICOR Hungary <a.steiner@medicor.hu> ezt írta (időpont: 2021. okt. 7., Cs, 15:50):

Kedves Jutas!

Az oxigén szenzor beszállítója ismét jelezte felénk, hogy a KE-25 szenzor orvosi felhasználásra nincs engedélyeztetve.

"Please note this sensor is not approved as a medical product, it is only suitable for non-medical applications."

Van a Viamednek is más terméke, de valójában ez egy Figaro gyártmány (figaro.co.jp/en), náluk lehetne érdeklődni, hogy mit ajánlanak (figaro@figaro.co.jp). Sajnos a szenzor átmérője nem valószínű, hogy ugyanaz lesz.

Vagy maradunk a KE-25-nél CE nélkül.

Üdv.

András



MEDICOR
Superbrands

András Arnold STEINER
Deputy CEO / Vice President

H-1097 Budapest, Illatos út 9., Hungary
Mobile: (+36 30) 921 0672
Phone: (+36 1) 280 6900
Fax: (+36 1) 280 6336
www.medicor.hu

MEDICOR ELEKTRONIKA ZRT.
MEDICAL EQUIPMENT AND DEVICES FOR NEONATES

BABYLIFE[®] Selection CE 2409

----- Továbbított üzenet -----

Tárgy:Re: New order 06OCT2021

Dátum:Thu, 7 Oct 2021 09:02:54 +0100

Feladó:Zoey Teal <office@viamed.co.uk>

Válaszcím:zoey.teal@viamed.co.uk

Címzett:Mr. Andras Arnold STEINER / MEDICOR Hungary <a.steiner@medicor.hu>

Dear Andras,

Thank you for your order.

The lead time will be approximately 3-5 weeks. Would you like me to go ahead and process your order?

Please note this sensor is not approved as a medical product, it is only suitable for non-medical applications.

If you have any further queries, please do not hesitate to contact me.

Kind regards

Zoey Teal

Please note: My working days are Monday to Thursday, 7.45am-6.00pm.

We are currently working remotely however telephone calls to the main office will still be answered.

<http://www.viamed.co.uk>

Email: zoey.teal@viamed.co.uk

Alternate email: viamed.zoey.teal@gmail.com

Tel: +44 (0)1535 634542

Fax: +44 (0)1535 635582

Sign up to our mailing list, go to www.viamed.co.uk/sign-up

Twitter: twitter.com/ViamedLtd

Facebook Page: facebook.com/ViamedLtdUK

Company registered in England, No. 01291765.

E-mail Disclaimer

GDPR / Privacy Policy available at www.viamed.co.uk/policy

The information in this email is confidential and may be legally privileged.

It is intended solely for the addressee. Access to this email by anyone else is unauthorized.

If you are not the intended recipient, any disclosure, copying, distribution or any action taken or omitted to be taken in reliance on it, is prohibited and may be unlawful.

If you feel you have received this email in error or would like to be removed from our mailing list, please reply with delete in the subject line or call +44 (0)1535 634542

--

 <p>MEDICOR Superbrands ce ce</p>	<p>Jutás Deák Head of R&D H-1097 Budapest, Illatos út 9., Hungary Mobile: (+36 20) 249 5682 Tel.: (1) 280 6900 / 110 Fax: (1) 280 6336 www.medicor.hu</p>	<p>MEDICOR ELEKTRONIKA ZRT. MEDICAL EQUIPMENT AND DEVICES FOR NEONATES</p>  
---	--	---