

Service Repair Sheet SRS67971

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Anthony Loughman

Royal Preston Hospital

Medical Engineering

01772520697

00004260

anthony.loughman@lthtr.nhs.uk

18/Oct/2021

Robert Connor

Viamed

Warranty

Notes 06/Oct/2021 Kate Griffiths

06/Oct/2021 Kate Griffiths

Out of box failure of R-17MED s/n 617289 to be sent in for warranty replacement

18/Oct/2021 Robert Connor

Received 1 x R17MED s/n 617289

Ready For quote

Repair Complete Signed

CGreen 18.10.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34289	Oxygen Sensor	0110017	617289	Yes

0140013 x 1 @ FOC  
old S/N, SRS, SRN

0110017 x 1 FOC  
New S/N, SRS, SRN

UPS x 1 @ £0.

Please find enclosed O2 sensor SN:617289  
Returns reference: SRS67971

*O2 Cell intermittent.*

From and to be Returned to:

Anthony Loughman  
Medical Engineering Department  
Royal Preston Hospital  
Sharoe Green Lane  
Preston  
PR2 9HT

Phone: 01772 524646

Email: [anthony.loughman@lthtr.nhs.uk](mailto:anthony.loughman@lthtr.nhs.uk)

---