

Service Repair Sheet SRS67946

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Zainab Jaffer

Barnet Hospital MES Dept

Equipment Librarian

0208 2164978

00000261

zainab.jaffer.ext@siemens-healthineers.com

08/Sep/2021

Robert Connor

Viamed

Quote

Notes 08/Sep/2021 Robert Connor

08/Sep/2021 Robert Connor

1 x V1000 s/n PR01866A10, battery contacts corroded from water ingress

Ready For quote

CGreen 9-9-2021

Repair Complete Signed

CGreen 20-10-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34185	Foetal Simulator	1410000	PR01866A10	N

1480000 x 1 E4S
SIN, SRS, SRN.

1430309 x 1 FOC
SRS, SRN

1430288 x 1 E2-50
SRS, SRN

UPS x 1 E10

Official Purchase Order

Order no	35455712-A	PORF ref	BAR16461
Order date	15/10/2021	Terms	60 days

Contact	Authorisation
Atyza Mamdouh Tel: +44 (0) 208 216 4978 barnet-admin.med.gb@siemens-healthineers.com	15/10/2021 X <i>De Silva</i> Signed by: De Silva Edrimun

Billing address	Delivery address
Siemens Healthcare Ltd c/o Data Capture Solutions P.O. Box 4728, Slough SL1 0AE Tel: +44 (0) 845 602 6873 InvoicesHealthcare.GB@siemens.com	Barnet Hospital Siemens Medical Equipment Services Level 0, Phase 1B Wellhouse Lane Barnet, Hertfordshire EN5 3DJ

Supplier
VIAMED LTD 15 STATION ROAD, CROSSHILLS KEIGHLEY WEST YORKSHIRE BD20 7DT Tel: 0153 563 4542 Fax: 0153 563 5582 Email: orders@viamed.co.uk

Type of order	Repair order	Requestor	☒
---------------	--------------	-----------	-------------------------------------

Part no	Description	Job no	Cost	Qty	Total
PRO1866A10	Foetal Heart Simulator V1000	433981	£ 45.00	1	£ 45.00
	V1000 Battery Clip 4 x AA	433981	£ 2.50	1	£ 2.50
Carriage					£ 10.00
Total excluding VAT					£ 57.50