



VIA MED

Resuscitation Cabinet

Class I
Via Rule Rule 9
Assesment Route Annex II
NBOG MD 1100

Carried out by Derek Lamb
21 / 10 / 21



VIA MED

-  Stock References Review
-  Supplier Review
-  Sales Review
-  Countries Review
-  Returns Review
-  Design Changes Review
-  User Instructions Review
-  Labels Review
-  Document Updates Review
-  Internal Issues Review
-  Clinical / FDA Incidents online search

Stock References Review

Stock Reference	Description
0120140	Adapter 15mm I.D. - 15mm I.D
0310001	Infant resuscitation cabinet.
0310002	Infant resuscitation cabinet.
0310003	Wall mounted cabinet including :
0310004	Cabinet - Type 2 no suction.
0310007	Infant resuscitation cabinet.
0310008	Infant resuscitation cabinet.
0310012	Infant resuscitation cabinet.
0310013	Infant resuscitation cabinet.
0310020	Storage bins size 3 blue.
0310025	Radiant warmer with double hinged arm
0310037	Rail mounted low suction controller
0310051	Infant resuscitation cabinet.
0310052	Infant resuscitation cabinet.
0310054	Infant resuscitation cabinet.
0310055	Infant resuscitation cabinet.
0310056	Infant resuscitation cabinet.
0310302	Radiant warmer with double hinged arm
0320010	Universal rail clamp with `V`
0320021	Resuscitation Bed 14 flat.`
0320023	Resuscitation Bed rotating.
0320039	Mattress octaganol - infant
0320043	Mattress large - infant resuscitation.
0320200	Mounting bracket - Medirail. For use
0330005	Resuscitation Cabinet Type 1
0330006	Door pad - for use with resus cabinet
0330007	Resuscitation Cabinet Door Hinge RHS.
0330008	Resuscitation Cabinet Door Hinge LHS.
0330009	Resuscitation Cabinet Door Catch
0330035	Suction clear tubing (1m)
0330105	Resuscitation Cabinet Type 2
0330106	Resuscitation Cabinet Type 3.
0332037	Resuscitation Cabinet Rail - Spacer 15mm
0332038	Resuscitation Cabinet Rail - Spacer 15mm
0332039	Resuscitation Cabinet Rail - Spacer 20mm
0332040	Resuscitation Cabinet Rail - Spacer 20mm
0332041	Resuscitation Cabinet Rail - Screw 35mm
0332042	Resuscitation Cabinet Rail - Screw 35mm
0332043	Resuscitation Cabinet Rail - Lock nut
0332044	Resuscitation Cabinet Rail - Lock nut
0332045	Resuscitation Cabinet Rail - Washer
0332046	Resuscitation Cabinet Rail - Washer

Stock Reference	Description
0332055	Resuscitation Cabinet Rail - 0.3m length
0332056	Resuscitation Cabinet Rail - 0.4m length
0332057	Resuscitation Cabinet Rail - 0.4m length
0332058	Resuscitation Cabinet Rail - 0.3m length
0332059	Resuscitation Cabinet Rail - 0.4m length
0332060	Resuscitation Cabinet MediRail ENDCAP
0332100	Resuscitation Cabinet Door Catch
0340010	Labour - Resuscitation Cabinet.
0340011	Training - Resuscitation Cabinet.
0350100	Packing box for resuscitation cabinet.
0380010	Annual service - Resuscitation cabinet.
0390000	B/W Resuscitation Cabinet Side 2
0390053	Label: Infant resuscitation cabinet.
0390100	Leaflet - Infant Resuscitation Cabinet.
0391055	Service sheet - Resuscitation Cabinet
3210071	Single use silicone round facemask
3833-007	Tapered connectors 10 per pack.
3833-132	1l VacSax canister with V mount.
9910-340	1l VacSax standard bore liners.Box of 25
9910-340S	1l VacSax standard bore liner - Single
9950050	Battery 1.5V AA 2700 mAh.

Comments on Stock references review:

part number upto date

products no longer on the market

Supplier Review

Stock Ref.	Description	Supplier A/C	Supplier P/N	Supplier Name	Rating
0320021	Resuscitation Bed 14 fl				Needs Grading
3833-007	Tapered connectors 10 p				Needs Grading
9950051	Batteries 1.5V AA. Proce				Needs Grading
0310025	Radiant warmer with doub	00007785	521A-60022OND	Nufer Medical AG	B
0310302	Radiant warmer with doub	00007785	521A-60022-3-EX	Nufer Medical AG	B
0320019	Resuscitation Bed 12 Fl	00009022	PP8720	Tanda Engineering Ltd	B
0320021	Resuscitation Bed 14 fl	00009022		Tanda Engineering Ltd	B
0320023	Resuscitation Bed rotat	00009022	PP8762	Tanda Engineering Ltd	B
0332055	Resuscitation Cabinet Ra	00009022	PP8765	Tanda Engineering Ltd	B
0332056	Resuscitation Cabinet Ra	00009022	PP8766	Tanda Engineering Ltd	B
0332059	Resuscitation Cabinet Ra	00009022	PP8768	Tanda Engineering Ltd	B
0332056	Resuscitation Cabinet Ra	00009032	PP8766	R & L Enterprises Ltd	B
0332058	Resuscitation Cabinet Ra	00009032	PP8767	R & L Enterprises Ltd	B
0332059	Resuscitation Cabinet Ra	00009032	PP8768	R & L Enterprises Ltd	B
0332037	Resuscitation Cabinet Ra	00009045		Essentra Components	B
0320043	Mattress large - infant	00009054	0320043	Angel Healthcare Solutio	D
0310020	Storage bins size 3 bl	00009061	2143013	Farnell	B
0330006	Door pad - for use with	00009061	1165061	Farnell	B
0332046	Resuscitation Cabinet Ra	00009061	1580684	Farnell	B
0360000	Hex wrench short arm 3	00009061	609-635	Farnell	B
9950051	Batteries 1.5V AA. Proce	00009061		Farnell	B
0350040	Patient circuit (single	00009082	INT-6181	Intersurgical Ltd	B
0360000	Hex wrench short arm 3	00009123	609-635	RS Components Ltd	B
0310037	Rail mounted low suction	00009150	4725-3M	Therapy Equipment Ltd	B
0320010	Universal rail clamp wit	00009150	2003	Therapy Equipment Ltd	B
0320030	1/2 litre suction jar an	00009150	4508	Therapy Equipment Ltd	B
0330034	Suction clear tubing (30	00009150	4302	Therapy Equipment Ltd	B
0330035	Suction clear tubing (1m	00009150	4302	Therapy Equipment Ltd	B
0330055	Universal clamp - 44mm h	00009150	2002	Therapy Equipment Ltd	B
0390000	B/W Resuscitation Cabine	00009360	0390000	Print Centre	C
0390100	Leaflet - Infant Resusci	00009360		Print Centre	C
0390100	Leaflet - Infant Resusci	00009363		ACI Print Group Impress	B
0332056	Resuscitation Cabinet Ra	00009365	PP8766	WEC Machining	B
0332058	Resuscitation Cabinet Ra	00009365	PP8767	WEC Machining	B
0332059	Resuscitation Cabinet Ra	00009365	PP8768	WEC Machining	B
0310035	Low suction controller	00009450		Oxylitre	B
0310141	Injector low suction con	00009450	S801/L	Oxylitre	B
0330040	Suction hose assembly (3	00009450		Oxylitre	B
3210071	Single use silicone roun	00009602	M6200500D	Bunzl Healthcare	B
3210071	Single use silicone roun	00009721	M6200500D	Marshall Products Limite	B
0330238	Rectangular Bracket (G+R	00010160	0330238	KSJ Engineering Limited	B
0330005	Resuscitation Cabinet T	00010220	BA/10 TYPE 1	Bristol Maid	B

Stock Ref.	Description	Supplier A/C	Supplier P/N	Supplier Name	Rating
0330105	Resuscitation Cabinet Ty	00010220	SPRC11	Bristol Maid	B
0390053	Label: Infant resuscitat	00010665		Pendle Signs & Plastics	B
0332040	Resuscitation Cabinet Ra	00011727	1111630000VR	Plastic Parts Centre Ltd	B
3833-007	Tapered connectors 10 p	00011810		Vacsax Ltd.	B
3833-132	1l VacSax canister with	00011810	3833-132	Vacsax Ltd.	B
9910-340	1l VacSax standard bore	00011810	9910-340	Vacsax Ltd.	B
0350100	Packing box for resuscit	00011818		Macfarlane Group	A
0120140	Adapter 15mm I.D. - 15m	00011844	PF 1515	Instrumentation Industri	B
0320039	Mattress octaganol - in	00012180	0320039	Park House Healthcare Lt	C
0320042	Mattress small - infant	00012180		Park House Healthcare Lt	C
0320043	Mattress large - infant	00012180	0320043	Park House Healthcare Lt	C
9950051	Batteries 1.5V AA. Proce	00012264		H-Squared UK Electric Lt	B
0330009	Resuscitation Cabinet Do	00012266	B344-12	R P Hardware Ltd	B
0330105	Resuscitation Cabinet Ty	00012407	SPRC11	W Shuttleworth & Co	B
9950051	Batteries 1.5V AA. Proce	00070006			Needs Grading

Comments on Suppliers:

products no longer on the market

Sales Information

Stock Reference	Description	2015	2016	2017	2018	2019	2020	2021
0120140	Adapter 15mm I.D. - 15m	149	90	35	92			
0310001	Infant resuscitation cabinet.							
0310002	Infant resuscitation cabinet.	7	12	20	13			
0310003	Wall mounted cabinet including :							
0310004	Cabinet - Type 2 no suction.							
0310007	Infant resuscitation cabinet.							
0310008	Infant resuscitation cabinet.							
0310012	Infant resuscitation cabinet.							
0310013	Infant resuscitation cabinet.							
0310020	Storage bins size 3 bl					5		
0310025	Radiant warmer with doub							
0310037	Rail mounted low suction							
0310051	Infant resuscitation cabinet.							
0310052	Infant resuscitation cabinet.							
0310054	Infant resuscitation cabinet.							
0310055	Infant resuscitation cabinet.							
0310056	Infant resuscitation cabinet.							
0310302	Radiant warmer with doub							
0320010	Universal rail clamp wit				10			
0320021	Resuscitation Bed 14 fl	9	20	18	8	13	6	14
0320023	Resuscitation Bed rotat							
0320039	Mattress octaganol - in							
0320043	Mattress large - infant	5	2	10	11	5	3	3
0320200	Mounting bracket - Medirail. For use	5			1			
0330005	Resuscitation Cabinet T							
0330006	Door pad - for use with							
0330007	Resuscitation Cabinet Door Hinge RHS.							
0330008	Resuscitation Cabinet Door Hinge LHS.							
0330009	Resuscitation Cabinet Do							
0330035	Suction clear tubing (1m	6		1	6	9	6	32
0330105	Resuscitation Cabinet Ty	5	1	2	4	1		
0330106	Resuscitation Cabinet Type 3.							
0332037	Resuscitation Cabinet Ra							
0332038	Resuscitation Cabinet Rail - Spacer 15mm							
0332039	Resuscitation Cabinet Rail - Spacer 20mm	19		16	32			
0332040	Resuscitation Cabinet Ra							
0332041	Resuscitation Cabinet Rail - Screw 35mm	19		16	32			
0332042	Resuscitation Cabinet Rail - Screw 35mm							
0332043	Resuscitation Cabinet Rail - Lock nut	19		16	32			
0332044	Resuscitation Cabinet Rail - Lock nut							
0332045	Resuscitation Cabinet Rail - Washer	19		16	32			
0332046	Resuscitation Cabinet Ra							

Stock Reference	Description	2015	2016	2017	2018	2019	2020	2021
0332055	Resuscitation Cabinet Ra							
0332056	Resuscitation Cabinet Ra							
0332057	Resuscitation Cabinet Rail - 0.4m length		60			10		1
0332058	Resuscitation Cabinet Ra	5		2	4			
0332059	Resuscitation Cabinet Ra	11		15	8			
0332060	Resuscitation Cabinet MediRail ENDCAP							
0332100	Resuscitation Cabinet Door Catch							
0340010	Labour - Resuscitation Cabinet.							
0340011	Training - Resuscitation Cabinet.							
0350100	Packing box for resuscit							
0380010	Annual service - Resuscitation cabinet.						4	17
0390000	B/W Resuscitation Cabine							
0390053	Label: Infant resuscitat							
0390100	Leaflet - Infant Resusci							
0391055	Service sheet - Resuscitation Cabinet							
3210071	Single use silicone roun	734	839	780	765	498	30	10
3833-007	Tapered connectors 10 p	1			5	5	8	5
3833-132	1l VacSax canister with				1	4	8	20
9910-340	1l VacSax standard bore	2			2	1		
9910-340S	1l VacSax standard bore liner - Single							
9950050	Battery 1.5V AA 2700 mAh.							

Comments on Sales Information:

only replacement parts being sold at this time,
main products have ceased

Countries Review

Country	2015	2016	2017	2018	2019	2020	2021
Canada	[X]						
Czech Republic							
France				[X]			
Germany						[X]	
Guernsey				[X]			
Ireland					[X]	[X]	
Italy		[X]		[X]			
Norway							
Scotland			[X]				
Sri Lanka	[X]						
Switzerland							
Turkey							
UK	[X]	[X]	[X]	[X]	[X]	[X]	[X]
UK United Kingdom	[X]	[X]					
United Kingdom	[X]	[X]					

Comments on Sales to Countries:

only selling replacement parts

Comments on Risks with Sales to Countries:

no risks

Returns and Q.A. Fails Review

Stock Reference	Fault	2015	2016	2017	2018	2019	2020	2021
0310002	Needs Evaluating							
0310025	Control Board Replaced							
0310025	No Fault Found			1				
0310025	No Fault Return To Stock			14				
0310025	Tested							
0310302	Burn Marks on Connector					1		
0310302	Not Tested			1				
0320021	Chipped			2				
0320021	Unchecked - Returned To Stock							
0320043	Unchecked - Returned To Stock							

Comments on Returns:

no returns

Comments on Risks with Returns and Potential Re-work:

no reworks

Design Changes Review

Showing Documents Filed in Y 14 Design Changes

Comments on Design Changes:

no changes

Comments on Risks with Design Changes:

no changes

User Instructions Review

Showing Documents Filed in F 5 User Instructions

Document ID	Description	Date Added/Updated
17738	Resuscitation Cabinet Instructions For Use (with Therapy Equipme...	24/10/16
17737	Resuscitation Cabinet Instructions For Use (with Oxylitre) v11.0	24/10/16
17736	Resuscitation Cabinet Operator Manual (without appedices)	24/10/16
14789	Instruction Manual - Air Oxygen Blenderl	19/02/15
14688	Ceratherm 300 Installation Instructions for modified casing part...	16/01/15
12878	0310140 - S8 series injector suction Operating Instructions	15/08/13
9241	Resuscitation Unit and TC400 Instructions for Use / Mobile Cerat...	18/10/11
8510	Resuscitation Unit and TC400 Instructions for Use User Manual O...	06/07/11
8509	Resuscitation Unit and TC400 Instructions for Use User Manual O...	06/07/11
8508	Resuscitation Unit and TC400 Instructions for Use User Manual T...	06/07/11
8507	Operators manual for Inspiration Blender Resuscitation Unit and ...	06/07/11
8179	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	24/03/11
8178	Resuscitation Unit Instructions for Use / Installation Ceratherm...	24/03/11
8176	Resuscitation Unit and TC400 Instructions for Use User Manual T...	24/03/11
8175	Resuscitation Unit and TC400 Instructions for Use User Manual A...	24/03/11
8170	Resuscitation Unit and TC400 Instructions for Use User Manual O...	24/03/11
8126	Resuscitation Unit and TC400 Instructions for Use TC400/W mkl	23/03/11
8124	Resuscitation Unit and TC400 Instructions for Use / User Manual	23/03/11
4128	Resuscitation Unit and TC400 Instructions for Use TC400 prevent...	09/07/08
4127	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4121	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4120	Resuscitation Unit and TC400 Instructions for Use / Therapy Equi...	09/07/08
4119	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4118	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4117	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4115	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4114	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4113	Resuscitation Unit and TC400 Instructions for Use / Ceratherm Ch...	09/07/08
4112	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
4101	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	09/07/08
3977	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
3976	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
3975	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
3974	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
3973	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
3972	Resuscitation Unit and TC400 Instructions for Use / User Manual ...	08/07/08
2573	ARCHIVED VERSION Resuscitation Unit Instructions for Use / User ...	17/08/07
1511	User Manual Apgar Smiths NS 22G CE FILE Apgar Timer V1r Instruct...	09/05/07
1312	Resuscitation Unit Instructions for Use / User Manual Nufer Wall...	19/03/07
1308	Resuscitation Unit Instructions for Use / Operator Manual Cerat...	19/03/07
1303	Resuscitation Unit Instructions for Use / User Manual Oxylitre S...	19/03/07

Comments on User Instructions:

no changes

Comments on Risks User Instructions:

no changes

Labels Review

Showing Documents Filed in F 7 / F 8 Labels

Document ID	Description	Date Added/Updated
15598	Resuscitation Unit and TC400 Labels (2)	07/09/15
8127	Resuscitation Unit and TC400 Labels (1)	23/03/11
4110	Resuscitation Unit and TC400 Labels Caution 2006	09/07/08
4109	Resuscitation Unit and TC400 Labels cabinet labels	09/07/08
4108	Resuscitation Unit and TC400 Labels 1989	09/07/08
3983	TC400 Rear Label original.pdf CE FILE Resuscitation Unit and TC4...	08/07/08
3981	Resuscitation Unit and TC400 Labels TC400 2001 (5).pdf	08/07/08
3980	Resuscitation Unit and TC400 Labels TC400 2001.pdf	08/07/08
3918	Resuscitation Unit Labels rotating bed	23/06/08
1319	Resuscitation Unit Labels Apgar Timer (2002)	19/03/07
1318	Resuscitation Unit Labels Thermacot 2	19/03/07
1317	Resuscitation Unit Labels Ceratherm 2006 Weee	19/03/07
1316	Resuscitation Unit Labels Thermacot	19/03/07
1315	Resuscitation Unit Labels Caution Label	19/03/07

Comments on Labels:

no changes

Comments on Risks Labels:

no changes

Documentation Updates / Changes

Document ID	Description	Date Added/Updated
65830	Resuscitation Unit EC Declaration of Conformance Nufer Ceratherm...	28/07/21
65826	Notified body certification TUV Nufer ISO 13485 : 2003 Nufer ISO...	28/07/21
53079	Service Manual Ceratherm 600-3 Radiant Warmer	04/02/21
43759	Resuscitation Unit and TC400 Clinical Trials Reports Reviews and...	16/09/20

Comments on Document Changes:

no changes

Comments on Risks with Document Changes:

no changes

Internal Issues Review

Number of Issues reviewed: 6

Issue ID	Subject
206796	Connectors not for sale in EU
114403	Leaflet Update Infant Resus Cabinet Leaflet Leaflet Designed
147848	Oxygen Monitoring Catalogue inc LF sensors: 1st Checking Stage
108651	Price List NeoPEEP Price List Price List Signed off by Director
162884	Revised Service Manual Ceratherm 600-3 v14.176.E.pdf
107601	Tom Thumb Review 8th November 2017 - Discontinued Variants

Comments on Issues:

nothing on concern found in the issue system

Comments on Risks with Issues:

no risks

Web Site Updates - Web Site Updates

Connectors not for sale in EU

Issue ID #206796

Date Created 03/Dec/2020

Issued To: Catrin Hird

Completed Status: 09 Dec 2020

Notes:

Added by Steve Hardaker sent to Catrin Hird

A customer tried to order some connectors that are not for sale in the EU, should it say so on the website or have you been asked not to state that? p/n 0120130, 0120131, 0120133

04 Dec 2020 Catrin Hird

Updated, thank you for noticing.

Also updated the front page descriptions for:

0120104

0120106

0120108

0120109

0120139

0120140

0120141

0120142

0120151

0120164

0120165

0120166

09 Dec 2020 Steve Hardaker

Issues:206796

Marketing Job Logger - Marketing Job Logger

Leaflet Update Infant Resus Cabinet Leaflet Leaflet Designed

Issue ID #114403

Date Created 13/Feb/2018

Issued To: Steve Hardaker

Completed Status: 21 Oct 2021

Notes:

Catrin Hird 13/02/2018 Issue Generated
Leaflet Update Infant Resus Cabinet Leaflet Leaflet Designed

13 Feb 2018 Catrin Hird

Next Action Changed From To Steve Hardaker
Priority Changed From 5 To 0 Hi,

I have been asked to amend the current infant resus cabinet leaflet so that it lists the individual parts that can be purchased...

As standard I've got:

Cabinet 0330105
Tom Thumb 0310030
Low suction controllers; oxytitre 0310140?
Therapy 7725 0310048 or 0310050
Cereatherm 0310302
Blender 0310200

But if we are offering a pick and mix sort of option then are there a large number of parts we can include?

Blenders - including precision or you happy to stick with Tom Thumb?

I guess we should also be including Apgar, AX300, Liners?

I think we have gone to look at this years ago but not sure where we got to.

15 Feb 2018 Steve Hardaker

Low suction controllers; oxytitre 0310140?
- This is not a standard unit, was used as a problem solver for 1 customer with no piped vacuum, offer 0310035 S714R with 3m hose instead.

Therapy 7725 0310048 or 0310050 - go with 0310050 which has a 3m hose but state shorter hoses available upon request.

Blenders - stick with Inspiration for now, I will discuss Precision Medical ones if the customer wants an alternative.

You can include the part numbers of the other main items, such as Apgar timer, Vacsax jars, liners, bed, mattress etc but leave off the AX300: we will offer it as a custom-requirement if it comes up but only 1 customer ever had them and took them all out again immediately.

The complete parts list can be found in the user instructions, you don't want to add absolutely everything as it goes down to spare parts, but the part numbers for the main items are all in there. Note: the version on L is out of date, the latest (v11.0) is attached to Issue 70173 for SN to upload.

16 Feb 2018 Catrin Hird
Fab thanks

19 Feb 2018 Steve Hardaker

Issues:114403

Marketing Job Logger - Marketing Job Logger

Oxygen Monitoring Catalogue inc LF sensors: 1st Checking Stage

Issue ID #147848

Date Created 20/Jun/2019

Issued To: General

Completed Status: 21 Oct 2021

Linked Files

28 Aug 2019 Oxygen Monitoring Catalogue 2019-08-28 proof1

09 Sep 2019 Oxygen Monitoring Catalogue 2019-09-09 proof2

09 Sep 2019 Oxygen Monitoring Catalogue 2019-09-09 proof3

09 Sep 2019 Oxygen Monitoring Catalogue Signed off by SJH 2019-09-09

09 Oct 2019 Oxygen Monitoring Autumn 2019 2019-10-09 v1

Notes:

Catrin Hird 20/06/2019 Issue Generated

Leaflet Leaflet/Catalogue: Lead-free sensors 1st Checking Stage

28 Aug 2019 Catrin Hird

Next Action Changed From To Steve Hardaker

Subject Changed From Leaflet Leaflet/Catalogue: Lead-free sensors 1st Checking Stage

Subject Changed To Oxygen Monitoring Catalogue inc LF sensors: 1st Checking Stage Please find attached latest copy of the Oxygen Monitoring Catalogue.

All information is correct apart from awaiting text from Steve regarding the Benefits for the lead-free for each manufacturer.

Hard copy with proof sheet provided.

05 Sep 2019 Steve Hardaker

Next Action Changed From Steve Hardaker To Catrin Hird p1 - SN suggests trying different design style as the white box looks `stuck on`, and to my eyes its not really aligned with the other elements on the page. Is the white background fixed? if not, perhaps try the lighter blue?

p2 - standardise on lower case for oxygen sensors.

p5 - UFO-130-1 & UFO-130-2 datasheet says 3.4V +/- 0.4V at 100% oxygen, I think we need that last bit and it looks like we've always specified this incorrectly.

p8 - 2nd to last paragraph: add comma, change `of` to `for` so it reads "in preparation for this,"

p10 - 0110707 - missing image

p11 - SN has suggested 2 Envitec tagline bulletpoints, added to draft.

Note: I have checked the specs against the datasheets but have only checked the Envitec lead-free range for 0110800 - 0110808 as we don't seem to have the rest of the spec sheets on L drive. Have asked SN if he can get them and add to L.

One final thing, I'm not sure the "incorporating lead-free" logo on the front works, it seems like it should clearly state "range now incorporates lead-free sensors" or similar, but that would involve a minor re-design to that bottom corner. Perhaps one for discussion with you, SN and me next week.

Have put back on your desk with proofing sheet.

09 Sep 2019 Catrin Hird

Next Action Changed From Catrin Hird To Steve Hardaker Sorry there is a mix up, you have put notes for the sensor leaflet on the catalogue issue. Have updated 147847.

This issue is for the catalogue.

09 Sep 2019 Catrin Hird
Updated as per leaflet

09 Sep 2019 Steve Hardaker

Next Action Changed From Steve Hardaker To Catrin Hird p10 & 11 - Can you fit "@ 100% oxygen" directly after the output spec? The footer on p11 seems too remote.
p32 - 0120127 - Description incorrect, check the I.D. / O.D. details.
p33 - 0120140 - Not the correct part number for this item.
p38 - Image is black & white
p39 - Change General Purpose to lower case, 2 instances

09 Sep 2019 Catrin Hird

Next Action Changed From Catrin Hird To Steve Hardaker Updated.

Image for flowsensors is b& due to flowsensor leaflet, SN wanted the images to be similar rather than one colour and one not..

09 Sep 2019 Steve Hardaker

Suggested changing to colour images of flowsensors, unless there was a valid reason for using the black and white ones. Otherwise, no amends.

08 Oct 2019 Catrin Hird

Next Action Changed From Steve Hardaker To Steve Nixon Updated images.

Also amended catalogue title so reflects colour change on the leaflet version.

Attached to issue and provided to SN for sign off.

Steve N. Please proof and sign off asap, required for Medica.

09 Oct 2019 Catrin Hird
Redesigned front cover.

Reorder and remove some information from cross reference page (rename).

Hard copy with proof sheet provided to SN. Attached.

11 Oct 2019 Catrin Hird

Next Action Changed From Steve Nixon To General Signed off and added to L

Issues:147848

Marketing Job Logger - Marketing Job Logger

Price List NeoPEEP Price List Price List Signed off by Director

Issue ID #108651

Date Created 21/Nov/2017

Issued To: Steve Nixon

Completed Status: 21 Oct 2021

Linked Files

21 Nov 2017 NeoPEEP Price List 2017-11-21

21 Nov 2017 NeoPEEP Price List 2017-11-21 B

27 Nov 2017 NeoPEEP Price List 2017-11-27

28 Nov 2017 NeoPEEP Price List 2017-11-28

26 Mar 2018 NeoPEEP Price List 2018-03-26

Notes:

Catrin Hird 21/11/2017 Issue Generated

Price List NeoPEEP Price List Price List Signed off by Director

21 Nov 2017 Catrin Hird

Next Action Changed From To Katie Evans UK Price list generated and attached and provided to KE for review (copy also saved in Z Checking).

SN - prices missing for following parts and currencies:

0120151 0003 International End User Euros
0120151 0005 International End User USD
0120153 0002 UK Distributor
0120153 0003 International End User Euros
0120153 0004 International Distributor Euros
0120153 0005 International End User USD
3210013 0003 International End User Euros
3210013 0005 International End User USD
3210023 0003 International End User Euros
3210023 0005 International End User USD
3210067 0002 UK Distributor
3210067 0003 International End User Euros
3210067 0005 International End User USD
3210067 0006 International Distributor USD
3210068 0002 UK Distributor
3210068 0003 International End User Euros
3210068 0004 International Distributor Euros
3210068 0005 International End User USD
3210068 0006 International Distributor USD
3210069 0002 UK Distributor
3210069 0003 International End User Euros
3210069 0004 International Distributor Euros
3210069 0005 International End User USD
3210069 0006 International Distributor USD
3210070 0003 International End User Euros
3210070 0005 International End User USD
3210071 0003 International End User Euros
3210071 0005 International End User USD
3210073 0003 International End User Euros
3210073 0005 International End User USD
3210074 0003 International End User Euros

3210074 0005 International End User USD
3210076 0003 International End User Euros
3210076 0005 International End User USD
3210077 0003 International End User Euros
3210077 0005 International End User USD

22 Nov 2017 Katie Evans
Checked GBP end user lists and returned to your desk

27 Nov 2017 Catrin Hird
Reviewed and updated.
Hard copy of proof sheet provided.
Digital copy in Z:

28 Nov 2017 Catrin Hird

Next Action Changed From Katie Evans To Steve Nixon Updated and given to SN for review with proof sheet.

Also saved in Z:Information for LibraryIn Checking with SN

25 Jan 2018 Catrin Hird
GBP price list signed off. Above currencies need prices applying to be able to produce other price lists.

29 Jan 2018 Catrin Hird
The following prices are still missing:

0120153 0005 International End User USD
3210067 0002 UK Distributor
3210067 0006 International Distributor USD
3210068 0002 UK Distributor
3210068 0004 International Distributor Euros
3210068 0006 International Distributor USD
3210069 0002 UK Distributor
3210069 0004 International Distributor Euros
3210069 0006 International Distributor USD

12 Mar 2018 Catrin Hird
Above prices still missing

26 Mar 2018 Catrin Hird
Prices all been updated with Tom Thumb Price List.
Resaved into Z:Information for LibraryIn Checking with SNNeopeep Price Lists

Hard copies and proof sheet provided to SN, for checking. Zip file attached to this issue.

03 May 2018 Catrin Hird
SN signed off verbally. He adding to L drive.

Issues:108651

Intrastats - Information for Intrastats and L Drive

Revised Service Manual Ceratherm 600-3 v14.176.E.pdf

Issue ID #162884

Date Created 21/Jan/2020

Issued To: Steve Nixon

Completed Status: 23 Sep 2020

Linked Files

21 Jan 2020 Service Manual Ceratherm 600-3 v14.176.E.pdf

Notes:

Added by Steve Hardaker sent to Steve Nixon

Steve, please see attached an updated Service Manual Ceratherm 600-3 v14.176.E.pdf that I have just received from Daniel Urfer.

It includes a revised Technical check to remove the problematic power Level 1 current test.

I have added it to the Technical Files and linked it to the stock records for 0310300, 0310301, 0310302, 0310303, 0310304 and 0310305.

Please can you add it to L or bounce this to Derek if that easier? Thanks.

Issues:162884

VIAMED Sales And Marketing - Sales

Tom Thumb Review 8th November 2017 - Discontinued Variants

Issue ID #107601

Date Created 08/Nov/2017

Issued To: Catrin Hird

Completed Status: 21 Oct 2021

Linked Files

20 Nov 2017 Infant Resuscitation Catalogue 2017-11-20

20 Nov 2017 Tom Thumb Price List 2017-11-20 DRAFT WEIGHTINGS ISSUE OUTSTANDING WITH DL

21 Nov 2017 Tom Thumb Price List 2017-11-21 DRAFT WEIGHTINGS ISSUE OUTSTANDING WITH DL

21 Nov 2017 Tom Thumb GBP Price List 2017-11-21

21 Nov 2017 Tom Thumb GBP Price List 2017-11-21 b

21 Nov 2017 Infant Resuscitation Catalogue 2017-11-21

21 Nov 2017 Infant Resuscitation Catalogue 2017-11-21 b

Notes:

Copy Issue Main Issue #107599

Added by Steve Nixon sent to Derek Lamb

As of today we will discontinuing non relevant Tom Thumb (Infant T-piece Resuscitator) devices, either due to using non blended gas, specials or low/no sales. See attached list for further details.

Variants retained & continued:

0310030 TT-480

Required for general sales and for use with resuscitation cabinets

0310094 TT-480 with pole mount

Continue, but monitor - low sales to date.

Catrin, please edit price lists, leaflet, catalogues & web site accordingly, i.e. just retain 0310030 TT-480

Once reviewed and appraised, may add 0310094.

20 Nov 2017 Catrin Hird

All variants of Tom thumb apart from 0310030, removed from website.

Infant Resus Catalogue - removed tom thumb variants and updated image - still needs work regarding acrylics being removed - attached current artwork for review by SN - i.e. page 9 has some white space left - would you like further information regarding specifications to be included (even though we are trying to steer away from that in catalogues)?!?

Amended pages 14/15 regarding "also available" pages.

Hard copy provided to SN with proofing sheet.

Price List

9901-22 shown on price list - is 0330191 correct replacement part? Have updated price list to generate from Intrastats - issue with DL regarding weightings not working (108584) but attached for reference.

The following prices as required to update other price list (if required)... Hard copy provided to SN with proofing sheet.

0310030 0005 International End User USD

0330191 0002 UK Distributor

0330191 0004 International Distributor Euros

0330191 0006 International Distributor USD

3210011 0003 International End User Euros

3210011 0005 International End User USD

3210013 0003 International End User Euros

3210013 0005 International End User USD
3210067 0002 UK Distributor
3210067 0003 International End User Euros
3210067 0005 International End User USD
3210067 0006 International Distributor USD
3210068 0002 UK Distributor
3210068 0003 International End User Euros
3210068 0004 International Distributor Euros
3210068 0005 International End User USD
3210068 0006 International Distributor USD
3210069 0002 UK Distributor
3210069 0003 International End User Euros
3210069 0004 International Distributor Euros
3210069 0005 International End User USD
3210069 0006 International Distributor USD
3210070 0003 International End User Euros
3210070 0005 International End User USD
3210071 0003 International End User Euros
3210071 0005 International End User USD
3210073 0003 International End User Euros
3210073 0005 International End User USD
3210074 0003 International End User Euros
3210074 0005 International End User USD
3210076 0003 International End User Euros
3210076 0005 International End User USD
3210077 0003 International End User Euros
3210077 0005 International End User USD

21 Nov 2017 Catrin Hird

9901-22/0330191 removed from price list as requested.

Incorrect part numbers given for silicone adapters, should have been 0120150 (instead of 0120149),
0120152 (instead of 0120159).

Updated prices required list...

0120151 0003 International End User Euros
0120151 0005 International End User USD
0120153 0002 UK Distributor
0120153 0003 International End User Euros
0120153 0004 International Distributor Euros
0120153 0005 International End User USD
0310030 0005 International End User USD
3210013 0003 International End User Euros
3210013 0005 International End User USD
3210023 0003 International End User Euros
3210023 0005 International End User USD
3210067 0002 UK Distributor
3210067 0003 International End User Euros
3210067 0005 International End User USD
3210067 0006 International Distributor USD
3210068 0002 UK Distributor
3210068 0003 International End User Euros
3210068 0004 International Distributor Euros
3210068 0005 International End User USD
3210068 0006 International Distributor USD
3210069 0002 UK Distributor
3210069 0003 International End User Euros
3210069 0004 International Distributor Euros
3210069 0005 International End User USD
3210069 0006 International Distributor USD
3210070 0003 International End User Euros
3210070 0005 International End User USD
3210071 0003 International End User Euros

3210071 0005 International End User USD
3210073 0003 International End User Euros
3210073 0005 International End User USD
3210074 0003 International End User Euros
3210074 0005 International End User USD
3210076 0003 International End User Euros
3210076 0005 International End User USD
3210077 0003 International End User Euros
3210077 0005 International End User USD

21 Nov 2017 Catrin Hird

Next Action Changed From Catrin Hird To Katie Evans Please review price list for correct descriptions.

Hard copy provided and saved in Z Checking, proof sheet provided.

21 Nov 2017 Catrin Hird
Updated and reattached.

21 Nov 2017 Catrin Hird
SN - Infant Resus catalogue provided to you for checking with proof sheet.

27 Nov 2017 Catrin Hird
Price list updated (still waiting for product weighing coding to be updated by DL #108584. Hard copy with proof sheet with KE.

28 Nov 2017 Catrin Hird
Created Related Issue #109225
Added by Catrin Hird sent to Steve Nixon
Price list updated and sent to SN for review with proof sheet. Pricing missing as follows:

0120151 0003 International End User Euros
0120151 0005 International End User USD
0120153 0002 UK Distributor
0120153 0003 International End User Euros
0120153 0004 International Distributor Euros
0120153 0005 International End User USD
0310030 0005 International End User USD
3210013 0003 International End User Euros
3210013 0005 International End User USD
3210023 0003 International End User Euros
3210023 0005 International End User USD
3210067 0002 UK Distributor
3210067 0003 International End User Euros
3210067 0005 International End User USD
3210067 0006 International Distributor USD
3210068 0002 UK Distributor
3210068 0003 International End User Euros
3210068 0004 International Distributor Euros
3210068 0005 International End User USD
3210068 0006 International Distributor USD
3210069 0002 UK Distributor
3210069 0003 International End User Euros
3210069 0004 International Distributor Euros
3210069 0005 International End User USD
3210069 0006 International Distributor USD
3210070 0003 International End User Euros
3210070 0005 International End User USD
3210071 0003 International End User Euros
3210071 0005 International End User USD
3210073 0003 International End User Euros
3210073 0005 International End User USD
3210074 0003 International End User Euros
3210074 0005 International End User USD

3210076 0003 International End User Euros
3210076 0005 International End User USD
3210077 0003 International End User Euros
3210077 0005 International End User USD

Also saved in: Z:Information for LibraryIn Checking with SN

28 Nov 2017 Catrin Hird

Next Action Changed From Katie Evans To Catrin Hird See related issue.

28 Nov 2017 Catrin Hird
see related issue

Issues:107601

Clinical / FDA Incidents online search

Clinical Investigation online review

Do any of the Results indicate a Risk / Problem : No

Do any of the Results indicate outdated Technology :

Comments on Clinical Search :

product discontinued

Review of online FDA Incident reports

Do any of the Results indicate a Risk / Problem : No

Do any of the Results indicate outdated Technology :

Comments on Clinical Search :

product discontinued