


## Modification to the gas hose access slot of the Viamed Infant Resuscitation Cabinet



### Notes:

These dimensions are the internal dimensions for the internal base of the cabinet.

The previous design of 4 gas access holes is to be replaced with a single slot.

The slot is to be 140 mm x 32 mm at the widest part, with the ends of the slot rounded off with a radius of 16 mm.

