




Department of Health and Social Security
Kingston By-Pass Road Surbiton Surrey

Telephone 01-398 4166 ext 208.



Mr J S Lamb
Viamed
15 Station Road
Cross Hills
Keighley
WEST YORKS BD20 7DT

Your reference

Our reference 06/02/05

Date

24th December 1985

Dear Mr Lamb


COT TOPS

We have recently heard from Stepping Hill Hospital that they have found the nine modified cot tops which you sent to them and that they are satisfactory.

We were happy with the modified sample which you sent to us but were waiting for the hospital's comments before we cleared this defect. Since hearing from them we have returned the original modified sample to the hospital for use in their maternity unit.

Thank you for your co-operation in this matter which has now been brought to a satisfactory conclusion.

Yours sincerely


C J GADD
STB1C
Hinchley Wood