

Gail Bell <viamed.gail.bell@gmail.com>

Fwd: MD300 with alarms

2 messages

Ryan Swaine <office@viamed.co.uk>

6 December 2021 at 12:49

Reply-To: ryan.swaine@vandagraph.co.uk

To: Gail Bell <gail.bell@vandagraph.co.uk>, Michael Lamb <michael.lamb@vandagraph.co.uk>

Please process the order for Marea Diving:

2x MD300 (these will need the R-17S 0110218 scanning to them) 6x R-22VAN

Thanks Ryan

Ryan Swaine VANDAGRAPH Ltd. http://www.vandagraph.co.uk Email ryan.swaine@vandagraph.co.uk Tel: +44 (0)1535 634900

Mob: +44 (0)7803 907117

----- Forwarded message ------

From: Francis Kalifa <mareadiving@wanadoo.fr>

Date: Sat, 4 Dec 2021 at 16:42 Subject: MD300 with alarms

To: Ryan Swaine <ryan.swaine@vandagraph.co.uk>

Hi Ryan

Have you available 2 units M300 with alarms?

If you could explain me something on this analyzer. We use it on Air and Oxygen panels and we have the alarms low and High on duty..

When the O2 percentage is wrong, the analyzer screams. I there a way to activate any system on this analyzer to shut up the alarms while the dive tech is adjusting the level of the percentage..

In fact, when some divers are working underwater, they are disturbed by the duration of this alarm..

Let me know...

In the meantime..

I need 2 units MD300 (without the wall mount) + 6 R22 (it's for Divesofts analyzers..)

Best regards

Francis

Marea

Professional Diving Equipment

RN 98 Les 4 Chemins

83130 La Garde

cell +33608059144

tel +33(0)494084094

Ryan Swaine <office@viamed.co.uk>

6 December 2021 at 12:52

Reply-To: ryan.swaine@vandagraph.co.uk To: Francis Kalifa <mareadiving@wanadoo.fr>

Cc: Gail Bell <gail.bell@vandagraph.co.uk>, Michael Lamb <michael.lamb@vandagraph.co.uk>

Hi Francis

Thank you very much for your order, we have the stock available for immediate dispatch.

The alarms can be adjusted, but not below 18%. You can press the alarm button and this will silence the audible alarm, but only for 2 minutes. Unfortunately the devices were built to medical, not diving, standards.

Best regards

Ryan

Ryan Swaine VANDAGRAPH Ltd. http://www.vandagraph.co.uk Email ryan.swaine@vandagraph.co.uk

Tel: +44 (0)1535 634900 Mob: +44 (0)7803 907117

[Quoted text hidden]