

08 December 2021

Please find the below purchase order:

Model: MICROSTIM DB

Serial Number: M0005065

Fault: power knob has broken and would need calibration afterwards

Our PO: GLB412227

Price of Repair: £66.00

Please email repair updates to serviceadministration@olympus.co.uk

Collection/Return address:

Medical Physics International

Unit 9-11

Bolton Tech Exchange

Spa Road

Bolton

BL1 4AY

Invoicing/Billing address:

KeyMed (Medical & Industrial Equipment) Ltd

Stock Road

Southend on Sea

Essex

SS2 5QH

Kind regards

Paige Clow

Service Administrator – Customer Care