

Service Repair Sheet SRS68019

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Nick Kerson

Worthing Hospital

01243831643

00012604

uhsussex.ebme.srh@nhs.net

07/Dec/2021

Robert Connor

Viamed

Warranty

Notes 30/Nov/2021 Zoey Teal

30/Nov/2021 Zoey Teal

Re sending a V1000 for repair. Serial number PR0773A30. Unit was repaired this month but is not working again. Only switches on if the control knob is set a minimum.

07/Dec/2021 Robert Connor

V1000 s/n PR0773A30, with grey cover.

Ready For quote

Repair Complete Signed

CCNEED 14.12.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34374	Foetal Simulator	1410000	PR0773A30	Yes (r)

1480000 x 1 FOC.

covered by
warranty

1430310 x 1 FOC
SRS, SRN

UPS x 1 FOC

Western Sussex Hospitals



NHS Foundation Trust

VIAMED Ltd
15, STATION ROAD
CROSS HILLS
KEIGHLEY
WEST YORKSHIRE
BD20 7DT

Electro Biomedical Engineering (EBME)
St Richard's Hospital
Spitalfield Lane
Chichester
West Sussex
PO19 6SE
Tel: 01243 831643
Fax: 01243 831406
ebme@wsht.nhs.uk
www.westernsussexhospitals.nhs.uk



30/11/2021

Dear Phil,

V1000 Foetal Heart Simulator SN:S/N: PR0773A30 , Your Ref: SRS68091.

As per our telephone conversation, please find enclosed our Foetal Heart Simulator V1000 for Investigation and repair

Yours Sincerely

Nick Kerson

EBME Specialist Technician

Uhsussex.ebme.srh@nhs.net