

Service Repair Sheet SRS68019

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Nick Kerson

Worthing Hospital

01243831643

00012604

uhsussex.ebme.srh@nhs.net

07/Dec/2021

Robert Connor

Viamed

Warranty

Notes 30/Nov/2021 Zoey Teal

30/Nov/2021 Zoey Teal

Re sending a V1000 for repair. Serial number PR0773A30. Unit was repaired this month but is not working again. Only switches on if the control knob is set a minimum.

07/Dec/2021 Robert Connor

V1000 s/n PR0773A30, with grey cover.

Ready For quote

Repair Complete Signed

CCNEEN 14.12.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34374	Foetal Simulator	1410000	PR0773A30	Yes (r)

14 80000 x 1 FOC.

covered by
warranty

14 30310 x 4 FOC edited 17.12.2021.
SRS, SRN

UPS x 1 FOC