

Service Repair Sheet SRS68028

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Derek Nairns

Company/ Hospital Name

Kettering General Hospital

Department

Medical Electronics

Position

Direct Phone

01536492269

General Phone

01536 492000

Opera Account

00002380

Email

d.nairns@nhs.net

Order Number

Date Received

20/Dec/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 14/Dec/2021 Zoey Teal

14/Dec/2021 Zoey Teal

Out of box failure. Serial numbers V108099 and V108087 bought in September 2021.

20/Dec/2021 Robert Connor

R43V pair s/n V108099/87

Ready For quote

Repair Complete Signed

CGNQQN 21-12-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34489	Oxygen Sensor	0110043	V108087	Yes
SRN34490	Oxygen Sensor	0110043	V108099	Yes

3 PAIR

0140013 x 1 FOC
old S/N's SRS, SRN

0110043 x 1 FOC
new S/N, SRS

UPS x 1 FOC

Kettering General Hospital

NHS Foundation Trust

Medical Devices Department
Rothwell Road
Kettering
Northants
NN16 8UZ

Direct Dial: 01536 492269
Direct Fax: 01536 492268

Web: www.kgh.nhs.uk

With compliments

Dear sir
Please find enclosed
O2 cells for checking
Ref .SRS.68028.
Sn.V108099
Sn.V108087
Fault-wouldn't calibrate when fitted .
If found faulty can they be replaced please
Many thanks

REGARDS



Derek Nairns
D.NAIRNS@NHS.NET