Service Repair Sheet SRS68027

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Ashley Davies

Company/ Hospital Name

Thames Medical

Department

Position

Office Manager

Direct Phone

General Phone

Opera Account

Email

01903522911

15/Dec/2021

00011982

ashley@thamesmedical.com

Order Number

Date Received

Booked in By

Main Company

Robert Connor Viamed

Type Return

Quote

Notes 08/Dec/2021 Kate Griffiths

08/Dec/2021 Kate Griffiths

Customer to send in 4420507, faulty IRMA. s/n IR00007142. "It is currently completely unresponsive but very new" 15/Dec/2021 Robert Connor

VM-2500M-Vet unverified s/n 81030153, with silicone cover, li-poly battery, IRMA s/n IR00007142, lemo adapter, and universal mounting plate. Capnograph serial no is inaccessible.

Ready For quote

Repair Complete Signed

NOON 22.12.2021

SRN	Equipment	Stock Ref	Serial Number	Warranty	
SRN34388	Capnograph	4410530	81030153	7	
SRN34389	Accessory	4420507	IR00007142	Yes	

4440000 ×1 0 £0 NO fault found SIN, SRS, SRN.

4440010 × 1 2 £0 SIN, SRS, SRN

Thames Medical Ltd

8 A'Becket Gardens

Worthing, West Sussex BN13 2BW

www.thamesmedical.com

email

info@thamesmedical.com

tel

01903 522911 01903 522922

fax 01903 522922

Experts in patient monitoring and anaesthesia

Thames Medical

PURCHASE ORDER

order#

PO-01682

date

09/12/2021

vendor address Viamed Ltd

15 Station Road Cross Hills

Keighley, West Yorkshire

BD20 7DT

contact

Kate

phone

01535 634542

fax

01535 635582

carrier

payment terms

UPS

Net 30

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2500 M1-RL2WR	VM2500M - Warranty Repair - Service Level 2 SN-81030153 IRMA SN-IR00007142	Warranty Repair of VM2500M Mainstream Capnograph	1	£0.00	Taxable	£0.00
remarks	SRS 68027 IRMA Completely unresponsive, new unit.Please advise before any work takes place.			any /	B-TOTAL	£0.00
	Crossways Vets RH204NA			TC	OTAL	£0.00