

BLUR - OX.

22/02/06

1) B.O.M

- a) Data sheets
- b) RoHS
- c) availability ~ cost


2) P.C.B

- a) manufacture
- b) 50, off 1000 off etc
- c) same + components

Circuit Any changes

- a) Layout.
- b) components

Software Any changes

- a) Logo Removed.
 - b) ~~Opto~~ Strategist O₂ Analysis 0-100. +
 - c) Nitrogen Analysis (Deduced)
 - d) Add alpha numerals eq. CAL. at 20.9 %.
 - e. Auto Switch off 5 min in 
- Location of USB skt. at Bottom.



P.H/ SSC

Design Review / R417 Annals

1/3/00

Add switch 2 ref voltage
Ref Vols. sent to check analogue 2 Degree
circuits
Can be programmed let to be periodic samples
Firmware is Firmware LCD or perhaps 0000
No words for Transmission problem
or use Facilitation symbol for Failure



P. Anderson