

Service Repair Sheet SRS68030

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Michael Taylor

County Hospital

Medical Electronics

Clinical Technologist

01785 230305

01785 257731

00004840

michael.taylor@uhnms.nhs.uk

20/Dec/2021

Robert Connor

Viamed

Quote

Notes 15/Dec/2021 Zoey Teal

15/Dec/2021 Zoey Teal

Sending in a V1000 for calibration. Serial number PR0495A12

20/Dec/2021 Robert Connor

1 x V1000 s/n PR0495A12, with 4 x AA battery and grey silicone cover.

Ready For quote

C.Green 21-12-2021

Repair Complete Signed

C.Green 6-1-2021

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34492	Foetal Simulator	1410000	PR0495A12	N

1480000 x 1 @ £45

S/N, SRS, SRN

1430309 x 1 £0

SRS, SRN

UPS x 1 £10

University Hospitals of North Midlands

OFFICIAL ORDER

Order Date : 23-12-2021

Order No : X014741

Must be quoted on all correspondence.

Deliver To:

RECEIPTS & DISTRIBUTION CENTRE COUNTY
RECEIPTS & DISTRIBUTION CENTRE COUNTY
COUNTY HOSPITAL
WESTON ROAD
STAFFORD
STAFFORDSHIRE
ST16 3SA

Requested delivery date: 23-12-2021

Invoice and Payment Enquiries To

UNIVERSITY HOSPITALS NORTH MIDLANDS
UNIVERSITY HOSPITALS NORTH MIDLANDS
PO Box 17394
Birmingham
Email: elfs.205uhn@cloud-trade.net
B9 9NN

All enquiries regarding this order to:

Contact : Jannine Heeley
Telephone :
Facsimile No. :
Email Address : jannine.heeley@uhn.nhs.uk

Supplier

Viaimed Ltd

Conditions

- This order is placed subject to the relevant NHS Terms and Conditions as detailed below (<https://www.gov.uk/government/publications/nhs-standard-terms-and-conditions-of-contract-for-the-purchase-of-goods-and-supply-of-services>)
 - Where a valid agreement exists for the items listed below the following NHS Terms and Conditions shall prevail (as applicable):
 - NHS Terms and Conditions for the Supply of Goods (Contract Version) Or NHS Terms and Conditions for the Provision of Services (Contract Version).
 - Where no valid agreement exists for the items listed below the following NHS Terms and Conditions shall prevail (as applicable):
 - NHS Terms and Conditions for the Supply of Goods (Purchase Order Version) Or NHS Terms and Conditions for the Provision of Services (Purchase Order Version).
- All goods must be delivered to the address stated (unless otherwise agreed in writing at the point of order) and must be accompanied by a delivery note quoting the above Purchase Order Number. Goods will only be accepted Mon-Thurs between 08am-4:30pm and Fri 8am-4pm (for Royal Stoke Warehouse and County Stores deliveries only).
- The above order number must be quoted on all advice notes, delivery notes, correspondence, invoices, acknowledgements etc.
- Any price, quantity or specification variances to that shown above must be notified immediately and agreed in writing prior to delivery otherwise delays can occur in the settlement of your Invoice.
- Invoices must be addressed to the details shown in above Invoice and Payment Enquiries section. Invoices must quote the above Order No. All suppliers must endeavour to submit invoices electronically to ELFs via Cloud Trade.

VAT Registration No: GB 654 932 808 EORI Code: GB 654 932 808 000

Goods or Services Required	Quantity	UOM	Contract Ref.	Unit Price	Line Value	VAT
not entered	45.00	EA		£1.00	£45.00	£9.00
CALIBRATE FOETAL HEART SIMULATOR SN PRO495A12						
CALIBRATE FOETAL HEART SIMULATOR SN PRO495A12						
not entered	10.00	EA		£1.00	£10.00	£2.00
DELIVERY						
DELIVERY						

Service Repair Sheet 68030

Contact Name	Michael Taylor
Company/ Hospital Name	County Hospital
Department	Medical Electronics
Position	Clinical Technologist
Direct Phone	01785 230305
General Phone	01785 257731
Opera Account	00004840
Email	michael.taylor@uhnm.nhs.uk
Order Number	X014741
Date Received	20/Dec/2021
Booked in By	Robert Connor

Repair	Ref	S/N	Equipment Type	
SRN34492	1410000	PR0495A12	Foetal Simulator	Complete - Repaired Time :0 Hour(s)

Parts Replaced

Qty: 1	1480000	Foetal Heart Simulator V1000 - Service
Qty: 1	1430309	V1000 Transducer interface cushion.

V1000 has had transducer interface cushion replaced and calibration has been checked
