

Service Repair Sheet SRS68017

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Constantin Ciurea

Company/ Hospital Name

SC Liamed SRL

Department

Position

Direct Phone

40268327490

General Phone

+40 02683274

Opera Account

00007594

Email

constantin.ciurea@liamed.ro

Order Number

Date Received

12/Jan/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

Notes 29/Nov/2021 Sarah Walton

29/Nov/2021 Sarah Walton

Faulty VM-2160 SN: T1901200065

They have tried 4 different sensors and in each case it does not measure SPO2 and then gives sensor fault.

12/Jan/2022 Robert Connor

1 x VM2165 s/n T1901200065, in silicone cover, 2 x 0014752 s/n CEK10180, FGK10050.

Ready For quote

CGNEAR 18.1.2022.

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34550	Pulse Oximeter	0012165	T1901200065	N
SRN34551	Pulse Oximetry Finger Sensor	0014752	CEK10180	N
SRN34552	Pulse Oximetry Finger Sensor	0014752	FGK10050	Yes

€

0042160 @ 0.25 @ €70 per hr = €17.50

S/N, SRS, SRN.

0012749. x 1 @ €25.00.
SRN, SRS (same as unit)

0040001 x 2 FOC - NO fault found
S/N's, SRS's, SRN's.

ins ?

UPS x 1
32 x 24 x 12cm.

Sent by

Company name: LIAMED SRL

Name/Department: Alex Cernat/Provision

Address: GRIVITEI A8

City/Postal Code: 500182

Country: ROMANIA

Ph/Fax/Telex no: +40268327490

V A T Registration no: RO10188824

**PROFORMA
INVOICE**

NO 8615071901

DATE:27.12.2021

Sent to :

Company name : Viamed Ltd

ACCOUNT NO.

Adress : 15 Station RoadCross HillsKeighley, West
Yorkshire BD20 7DT

City/Postal Code BD20 7DT

Country : Great Britain

Ph/Fax/Telex no : +44 (0) 1535 634542

V A T Registration no :

AWB NO

No of pieces..1

Total Gross Weight 1 KG

Total Net Weight 0.5 KG

Terms of delivery

CARRIER :

FULL DESCRIPTION OF GOODS	Customs Commodity Code	Country of origin	Qty	Unit Value and Currency	Sub Total Value & Currency
SPo2 sensor		Great Britain	2	10 eur	10 eur
TOTAL VALUE AND CURRENCY					10eur

MANUFACTURED BY: Company name: Viamed

Address:

City/Postal Code:

Country:

REASON FOR EXPORT : service purpose, no commercial value

I declare that above information is true and correct
to the best of my knowledge and belief

NAME

SIGNATURE

STAMP



Catherine Spence Required Reading 1 Calls 32 Issues 18 Unread 17 ISO Tasks |
 H/W: ML ZT Future Active Users Stock Search



Select Currency

Pounds | Euros | Usd

Repair SRS68017 Values in Pounds @ £1: £1.00

SRN34550	Pulse Oximeter	0012165			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		0012749	1	£0.00	£0.00
		0042160	0.25	£90.00	£22.50
				Parts Cost	£22.50

AS Res
Intrastats
CG 18-1-2022

£ 25.00.
€ 70.00 €17.50

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent