

Post market Surveillance Viamed Product T Adaptors
Carried out by Derek Lamb on 16 / 03 / 11

Comments on Stock References

No Comments

Comments on Suppliers

No Comments

Comments on Sales Information

No Comments

Comments on Sales Countrys

No Comments

Comments on Countrys Risks

No Comments

Comments on Stock Returns

Total Real Unit Failures rate 1.04% Cable Failures, 1981 Sold including cables sent with Units 11 reported failures 0.56% failure rate

Comments on Stock Return/ Re-Work Risks

Most Unit failures are a complete replacement. Cable Failure is a replacement.

Comments on Design Changes

See canadian.

Comments on Design Changes

See canadian.

Comments on Issues

No customer complaints or Feedback Mostly Sales or Return Issues

Comments on Risks from Issue

No risks identified