

Memo.

To : John Lamb.

From : Simon Watmough.

Date : 26-6-01.

Re : Cross connection of wiring within Nufer infant radiator.

Please see attached response from Mr Hartmann at Nufer.

I still remain worried about this issue.

If the heater was wired matching the brown & blue wiring to the mains inlet live & neutral, the live would be switched at the facia on/off switch and the neutral would be hard wired into the control panel. With the wiring apparently cross connected, the reverse is true, that the neutral is being switched with the live hard wired.

I consider this to be contrary to good electrical engineering practice and question Nufer's verdict. They seem to be more interested in the mains inlet filter than in the remainder of the circuitry. At the very least that the blue/brown external wiring is reversed internally to brown/blue.

I am not an expert with wiring conventions and certainly not with the internal design of the Nufer infant radiator however it seems that a fundamental error has been made and I would be very grateful if you would consider my views and make the final decision.

S Watmough.

cc. S Nixon.