

Service Repair Sheet SRS67980

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Ashley Davies

Thames Medical

01903 522911

00011982

asley@thamesmedical.com

18/Oct/2021

Robert Connor

Viamed

Quote

Quote  
QVM134005  
has been  
deleted cg 4.2.22

Notes 14/Oct/2021 Catherine Spence

14/Oct/2021 Catherine Spence

VM2500-S S/N: 80940160

SP02 and CO2 not working

18/Oct/2021 Robert Connor

Received 1 x VM2500S s/n 80940160, with battery, mounting plate, silicone cover, sampling line and clip sensor s/n ESHH00363.

Ready For quote

CGreen 22.11.2021

Repair Complete Signed

CGreen 2.4.2022.

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34295	Capnograph	4410540	80940160	2
SRN34296	Pulse Oximetry Ear Clip Sensor	0015000	ESHH00363	2

DIST

4440000 x 1 @ £90 EO  
S/N, SRS, SRN

Returning as un-repaired  
as Requested by Customer  
cg 4.2.22.

~~4420504 x 1 @ £807.50~~  
~~SRS, SRN~~

\* 0040001 x 1 @ 0  
S/N, SRS, SRN -> NO fault found.

UPS x 1 @ £10. EO..



Thames Medical Ltd  
Picket Gardens  
Burgess, West Sussex BN13 2BW  
www.thamesmedical.com  
email info@thamesmedical.com  
tel 01903 522911  
fax 01903 522922  
Experts in patient monitoring and anaesthesia

# PURCHASE ORDER

order # PO-01648  
date 15/10/2021

vendor address **Viamed Ltd**  
15 Station Road  
Cross Hills  
Keighley, West Yorkshire  
BD20 7DT

contact Kate  
phone 01535 634542  
fax 01535 635582

carrier

payment terms

Net 30

PS

ITEM	DESCRIPTION	VENDOR PRODUCT CODE	QUANTITY	UNIT PRICE	SPECIAL TAX RATE	SUB-TOTAL
TML-V-VM2500 S3-RL2WR	VM2500S - Warranty Repair - Service Level 2 SN-80940160	Warranty Repair of VM2500S Sidestream Capnograph	1	£0.00	Taxable	£0.00
<i>No warranty with OS. cg. 4-2-22.</i>						SUB-TOTAL £0.00
<i>Please quote before any work</i>						VAT £0.00
						TOTAL £0.00

Please acknowledge this order and advise the delivery date by fax : +44 (0) 1903 522922 and/or by email:  
info@thamesmedical.com

Service Repair Sheet 67980

Contact Name	Ashley Davies
Company/ Hospital Name	Thames Medical
General Phone	01903 522911
Opera Account	00011982
Email	asley@thamesmedical.com
Date Received	18/Oct/2021
Booked in By	Robert Connor

---

Repair	Ref	S/N	Equipment Type	
SRN34295	4410540	80940160	Capnograph	Return Unrepaired Time :0 Hour(s)

The Capnograph VM 2500-S Vet is being returned unrepaired as requested.

---

Repair	Ref	S/N	Equipment Type	
SRN34296	0015000	ESHH00363	Pulse Oximetry Ear Clip Sensor	No Fault Found Time :0 Hour(s)

The SpO2 Lingual Sensor has been tested. No fault was found.

---