

Service Repair Sheet SRS68013

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Sean Bunn

Company/ Hospital Name

Luton & Dunstable Hospital

Department

Clinical Engineering

Position

Direct Phone

01582718369

General Phone

01582 497042

Opera Account

00003490

Email

Sean.Bunn@ldh.nhs.uk

Order Number

Date Received

07/Dec/2021

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Quote

with
Supplier

Notes 24/Nov/2021 Zoey Teal

24/Nov/2021 Zoey Teal

Sending in a Capnograph to repair and recalibrate. It is showing Flash Error when turned on and the Gas Percentage is low when using the service menu.

Serial number: 80640160

07/Dec/2021 Robert Connor

1 x VM2500S sim 80640160, with silicone cover and 4 x AA battery

Ready For quote

CGN 8.2.2022

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34381	Capnograph	4410520	80640160	N

4440000 x 1 @ £90 per hr. Time: 1/4 hr @ £22.50.

4480500 x 1 @ £150
SRS, SRN

UPS x 1 £10

Reference: 114674



LUTON &
DUNSTABLE
UNIVERSITY
HOSPITAL

CLINICAL EXCELLENCE. QUALITY & SAFETY

Work Requested: Service + Repair
Flash Error

Date: 24/11/21

Hospital

Hospital	Luton & Dunstable University Hospital
Address	Lewsey Road, Luton, LU4 0DZ
Phone Number	01582 718 369
Email Address	CEWorkshop@ldh.nhs.uk

Device

Asset no	25310
Serial Number	80640160
Model	VM-2500-S
Description of device	Capnograph

Sender

Name	Sean Bunn
Signature	Sean
Position	Clinical Engineer

Receiver

FAO	Service department
Company	Unimed
Phone	N/A
Email	N/A

Catherine Spence Required Reading 1 Calls 23 Issues 14 Unread 13 ISO Tasks |
 H/W: KG Future Active Users

▼ Search



Select Currency

Pounds| Euros| Usd

Repair SRS68013 Values in Pounds @ £1: £1.00

SRN34381	Capnograph	4410520			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		4440000	0.25	£90.00	£22.50
		4480500	1	£150.00	£150.00
				Parts Cost	£172.50

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent