

Service Repair Sheet SRS68037

Date Promised to Return: 01 Jan 1970 -3600

Contact Name

Erik Sommer

Company/ Hospital Name

Hellmut Habel Gesellschaft m.b.H.

Department

Position

Direct Phone

436649657570

General Phone

Opera Account

00006343

Email

erik.sommer@habel-medizintechnik.at

Order Number

Date Received

01/Feb/2022

Booked in By

Robert Connor

Main Company

Viamed

Type Return

Warranty

Notes 06/Jan/2022 Sarah Walton

06/Jan/2022 Sarah Walton

From customer - We received the Display-Board with part number 0039030, but this one isn't working with the device. The defective Board in the device is following, after the start-up we receive an error message and the device isn't working.

S/N:T2006200029.

I have asked the customer to either just return the board or return the whole device back so that we can take a look.

01/Feb/2022 Robert Connor

Received 1 x VM-2160 s/n T2006200029

Ready For quote

CGNEER 8.2.2022

Repair Complete Signed

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34658	Pulse Oximeter	0012165	T2006200029	Yes

Dist.

0039031 x 1 @ €100.
SRS, SRN

0039027 x 1 @ €6.00
SRS, SRN

INS?

UPS x 1
~~32x24x12cm~~ @ 0.60kg 23x15x15cm

VIAMED

Return Order: SRS68037

15 Station Road, Cross Hills, Keighley
BD20 7DT WEST YORKSHIRE
ENGLAND

Ship & Invoice to:

Hellmut Habel Gesellschaft m.b.H.
Ignaz-Koeck-Strasse 20
1210 VIENNA
AUSTRIA / EUROPE
Tel.: + 43 / 1 / 292 66 42 - 0
Fax: + 43 / 1 / 290 12 45
einkauf@habel-medizintechnik.at

Enquiry no. L079251

Date: 18.01.2022

Your phone no. +44 1535 634542
Your fax no. +44 1535 635582
Your supplier no. 301782 GB287389593

DELIVERY DATE

Contact: Marion Baier
Our Customer No.: I0007025
Page: 1 / 1

Delivery terms: ex works

Shipment: UPS Standard

We herewith return for repair due to
reference number SRS68037

Part No.	Description	Qty.	Unit	Price	%	Total
99RETOURE	Viamed 1x VM-2160 Snr.: T2006200029	1,0	ea.	Lager 01		
	Failure: display defect			Lager 01		
	Return Nr. SRS68037			Lager 01		
				Lager 01		
	Value for custom: €285.-			Lager 01		
	*** Proforma Rechnung:L079251***			Lager 01		

www.habel-medizintechnik.at

FN: 125978m HG Wien, DVR Nr.: 0447501, UID: ATU14213509, ARA Nr.: 590
Hellmut Habel Gesellschaft m.b.H., A-1211 Wien, Ignaz Köck Straße 20
Tel.: +43/1/292 66 42-0, Fax: +43/1/290 12 45, e-mail: office@habel-medizintechnik.at

Catherine Spence Required Reading 1 Calls 23 Issues 14 Unread 13 ISO Tasks |
 H/W: KG Future Active Users Search



Select Currency

Pounds | Euros | Usd

Repair SRS68037 Values in Euros @ €1.20130: 1.00

SRN34658	Pulse Oximeter	0012165			
	Parts Required	Reference	Quantity	Unit Cost	Total Line
		0039031	1	€140.55	€140.55
		0039027	1	€9.31	€9.31
				Parts Cost	€149.86

AS Per Intra
 (g 4-2)
 €100.
 €6.

CHECK PRICES BEFORE SENDING QUOTE

Quote Sent