

Service Repair Sheet SRS68058

Contact Name

Company/ Hospital Name

Department

Position

Direct Phone

General Phone

Opera Account

Email

Order Number

Date Received

Booked in By

Main Company

Type Return

Date Promised to Return: 01 Jan 1970 -3600

Christopher Reid

BMI Blackheath Hospital

Materials Assistant

02083187722

07770 608384

00000555

christopher.reid@bmihealthcare.co.uk

01/Feb/2022

Robert Connor

Viamed

Quote

Notes 27/Jan/2022 Sarah Walton

27/Jan/2022 Sarah Walton

S/N: M0004992

Faulty DB3 unit. Going to return along with patient leads

01/Feb/2022 Robert Connor

Received 1 x Microstim DB3 s/n M0004992, with battery and patient leads.

Ready For quote

C Green 2.2.2022

Repair Complete Signed

C Green 11.2.2022

SRN	Equipment	Stock Ref	Serial Number	Warranty
SRN34664	Microstim DB3	2510000	M0004992	N

2540010 x 1 @ £45

S/N, SRS, SRN.

2520000 x 1 @ £0

SRS, SRN.

UPS x 1 @ £10

Purchase Order



Dispatched via Print

Purchase Order 023-067115	Date 03/02/2022	Revision	Page 1
Payment Terms 60 Days	Freight Terms Destination	Ship Via Common	
Buyer Christopher Reid	Phone 020 8318 7722	Currency GBP	

VIAMED LTD
15 Station Road
Cross Hills
KEIGHLEY BD20 7DT
United Kingdom

Deliver to: The Blackheath Hospital
40 - 42 Lee Terrace
Blackheath
London GT LON SE3 9UD
United Kingdom

Invoice To: BMI Business Services
Park Square
Bird Hall Lane CHEA SK3 0XN
United Kingdom

Vat Reg ID: GB235134980
Vendor: 102762

Please supply the following goods or services in accordance with the prevailing Purchase Order terms and conditions for BMI Healthcare which you hereby agree to be bound by and avail yourself of. Order number must appear on all invoices, delivery notes and correspondence.

Tax Exempt? N Tax exempt ID:		Replenishment Option: Standard						
Item / Description	PS Item Ref	Mfg ID	Quantity	Pk Size	Price	Extended Amt		
MAINT REPAIR - NURSING EQUIPMENT REPAIR OF MICROSTIM DB3 YOUR QUOTE QVM135250	KM1482		1	1	55.00	55.00		

Total VAT Code: S(20.00%) 11.00
Total PO Amount 66.00

Terms of Payment 90 days nett unless stated above.

Blackheath Hospital
The Blackheath Hospital
40 - 42 Lee Terrace
Blackheath
London GT LON SE3 9UD
United Kingdom

Authorised Signature

Service Repair Sheet 68058

Contact Name

Christopher Reid

Company/ Hospital Name

BMI Blackheath Hospital

Position

Materials Assistant

Direct Phone

02083187722

General Phone

07770 608384

Opera Account

00000555

Email

christopher.reid@bmihealthcare.co.uk

Order Number

023-067115

Date Received

01/Feb/2022

Booked in By

Robert Connor

Repair

Ref

S/N

Equipment Type

SRN34664

2510000

M0004992

Microstim DB3

Complete - Repaired
Time :0 Hour(s)

Parts Replaced

Qty: 1

2540010

Repair & Parts Fixed Charge -

Qty: 1

2520000

Microstim DB3 - Patient Lead.

The Microstim DB3 has been serviced and patient lead has been replaced.
